

**Digital Imaging and Communications in Medicine (DICOM)**

**Part 6: Data Dictionary**

*Published by*

**National Electrical Manufacturers Association**

1300 N. 17th Street

Rosslyn, Virginia 22209 USA

© Copyright 2007 by the National Electrical Manufacturers Association. All rights including translation into other languages, reserved under the Universal Copyright Convention, the Berne Convention for the Protection of Literacy and Artistic Works, and the International and Pan American Copyright Conventions.

## NOTICE AND DISCLAIMER

The information in this publication was considered technically sound by the consensus of persons engaged in the development and approval of the document at the time it was developed. Consensus does not necessarily mean that there is unanimous agreement among every person participating in the development of this document.

NEMA standards and guideline publications, of which the document contained herein is one, are developed through a voluntary consensus standards development process. This process brings together volunteers and/or seeks out the views of persons who have an interest in the topic covered by this publication. While NEMA administers the process and establishes rules to promote fairness in the development of consensus, it does not write the document and it does not independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in its standards and guideline publications.

NEMA disclaims liability for any personal injury, property, or other damages of any nature whatsoever, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. NEMA disclaims and makes no guaranty or warranty, expressed or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any of your particular purposes or needs. NEMA does not undertake to guarantee the performance of any individual manufacturer or seller's products or services by virtue of this standard or guide.

In publishing and making this document available, NEMA is not undertaking to render professional or other services for or on behalf of any person or entity, nor is NEMA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances. Information and other standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

NEMA has no power, nor does it undertake to police or enforce compliance with the contents of this document. NEMA does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health or safety-related information in this document shall not be attributable to NEMA and is solely the responsibility of the certifier or maker of the statement.

## CONTENTS

NOTICE AND DISCLAIMER .....	2
CONTENTS .....	3
FOREWORD .....	4
1 Scope and field of application .....	5
2 Normative references .....	5
3 Definitions .....	5
3.1 DICOM INTRODUCTION AND OVERVIEW DEFINITION .....	5
3.2 DICOM INFORMATION OBJECT DEFINITION .....	5
3.3 DICOM DATA STRUCTURES AND ENCODING DEFINITIONS .....	6
3.4 DICOM DATA DICTIONARY .....	6
4 Symbols and abbreviations .....	6
5 Conventions .....	6
6 Registry of DICOM data elements .....	8
7 Registry of DICOM File Meta Elements .....	84
8 Registry of DICOM directory structuring elements .....	85
Annex A Registry of DICOM unique identifiers (UID) (Normative) .....	86

## FOREWORD

The American College of Radiology (ACR) and the National Electrical Manufacturers Association (NEMA) formed a joint committee to develop a standard for Digital Imaging and Communications in Medicine (DICOM). This DICOM Standard was developed according to the NEMA procedures.

This standard is developed in liaison with other standardization organizations including CEN TC251 in Europe and JIRA in Japan, with review also by other organizations including IEEE, HL7 and ANSI in the USA.

The DICOM Standard is structured as a multi-part document using the guidelines established in the following document:

ISO/IEC Directives, 1989 Part 3 : Drafting and Presentation of International Standards.

This document is one part of the DICOM Standard which consists of the following parts:

- PS 3.1: Introduction and Overview
- PS 3.2: Conformance
- PS 3.3: Information Object Definitions
- PS 3.4: Service Class Specifications
- PS 3.5: Data Structures and Encoding
- PS 3.6: Data Dictionary
- PS 3.7: Message Exchange
- PS 3.8: Network Communication Support for Message Exchange
- PS 3.9: Retired
- PS 3.10: Media Storage and File Format for Media Interchange
- PS 3.11: Media Storage Application Profiles
- PS 3.12: Media Formats and Physical Media for Media Interchange
- PS 3.13: Retired
- PS 3.14: Grayscale Standard Display Function
- PS 3.15: Security and System Management Profiles
- PS 3.16: Content Mapping Resource
- PS 3.17: Explanatory Information
- PS 3.18: Web Access to DICOM Persistent Objects (WADO)

These parts are related but independent documents. Their development level and approval status may differ. Additional parts may be added to this multi-part standard. PS 3.1 should be used as the base reference for the current parts of this standard.

## 1 Scope and field of application

This part of the DICOM Standard is PS 3.6 of a multi-part standard produced to facilitate the interchange of information between digital imaging computer systems in medical environments. This interchange will enhance diagnostic imaging and potentially other clinical applications. The multi-part DICOM Standard covers the protocols and data that shall be supplied to achieve this interchange of information.

This part of the standard contains the registry of all DICOM Data Elements and all DICOM Unique Identifiers that are defined within the DICOM Standard.

## 2 Normative references

The following standards contain provisions that, through references in this text, constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this standard are encouraged to investigate the possibilities of applying the most recent editions of the standards indicated below.

ACR-NEMA 300-1988 Digital Imaging and Communications

ISO 8649:1988 Information processing systems - Open Systems Interconnection - Service definition for the Association Control Service Element (ACSE).

ISO 8822:1988 Information processing systems - Open Systems Interconnection - Connection oriented presentation service definition.

ISO/IEC Directives, 1989 Part 3 - Drafting and presentation of International Standards.

## 3 Definitions

For the purposes of this standard, the following definitions apply.

### 3.1 DICOM INTRODUCTION AND OVERVIEW DEFINITION

This part of the standard makes use of the following term defined in PS 3.1:

- Attribute

### 3.2 DICOM INFORMATION OBJECT DEFINITION

This part of the standard makes use of the following term defined in PS 3.3:

- Attribute Tag

### 3.3 DICOM DATA STRUCTURES AND ENCODING DEFINITIONS

This part of the standard makes use of the following terms defined in PS 3.5:

- a. Data Element
- b. Data Element Tag
- c. Element Number
- d. Group Number
- e. Repeating Group
- f. Retired Data Element
- g. Standard Data Element
- h. Value Multiplicity (VM)
- i. Value Representation (VR)

### 3.4 DICOM DATA DICTIONARY

The following definition is commonly used in this Standard:

**Tag:** A unique identifier for an element of information composed of an ordered pair of numbers (a Group Number followed by an Element Number), which is used to identify Attributes and corresponding Data Elements.

## 4 Symbols and abbreviations

The following symbols and abbreviations are used in this Standard.

<b>ACR</b>	American College of Radiology
<b>DICOM</b>	Digital Imaging and Communications in Medicine
<b>IOD</b>	Information Object Definition
<b>ISO</b>	International Standards Organization
<b>NEMA</b>	National Electrical Manufacturers Association
<b>OSI</b>	Open Systems Interconnection
<b>TCP/IP</b>	Transmission Control Protocol/Internet Protocol
<b>UID</b>	Unique Identifier
<b>VM</b>	Value Multiplicity
<b>VR</b>	Value Representation

## 5 Conventions

Word(s) are capitalized in this document to help the reader understand that these word(s) have been previously defined in Section 3 and are to be interpreted with that meaning.

A Data Element Tag is represented as (gggg,eeee), where gggg equates to the Group Number and eeee equates to the Element Number within that Group. Data Element Tags are represented in hexadecimal notation as specified for each named Data Element in this Standard.

“RET” is used to indicate that the corresponding Data Element, SOP Class, or Transfer Syntax has been retired. Retired items are shown italicized.

Note: The use of retired items is supported in this version of DICOM. However, new implementations are strongly encouraged to implement alternative Data Elements, SOP Classes or Transfer Syntaxes.

## 6 Registry of DICOM data elements

Note: For attributes that were present in ACR-NEMA 1.0 and 2.0 and that have been retired, the specifications of Value Representation and Value Multiplicity provided are recommendations for the purpose of interpreting their values in objects created in accordance with earlier versions of this standard. These recommendations are suggested as most appropriate for a particular attribute; however, there is no guarantee that historical objects will not violate some requirements or specified VR and/or VM.

Tag	Name	VR	VM	
(0008,0001)	<i>Length to End</i>	UL	1	RET
(0008,0005)	Specific Character Set	CS	1-n	
(0008,0008)	Image Type	CS	1-n	
(0008,0010)	<i>Recognition Code</i>	CS	1	RET
(0008,0012)	Instance Creation Date	DA	1	
(0008,0013)	Instance Creation Time	TM	1	
(0008,0014)	Instance Creator UID	UI	1	
(0008,0016)	SOP Class UID	UI	1	
(0008,0018)	SOP Instance UID	UI	1	
(0008,001A)	Related General SOP Class UID	UI	1-n	
(0008,001B)	Original Specialized SOP Class UID	UI	1	
(0008,0020)	Study Date	DA	1	
(0008,0021)	Series Date	DA	1	
(0008,0022)	Acquisition Date	DA	1	
(0008,0023)	Content Date	DA	1	
(0008,0024)	<i>Overlay Date</i>	DA	1	RET
(0008,0025)	<i>Curve Date</i>	DA	1	RET
(0008,002A)	Acquisition Datetime	DT	1	
(0008,0030)	Study Time	TM	1	
(0008,0031)	Series Time	TM	1	
(0008,0032)	Acquisition Time	TM	1	
(0008,0033)	Content Time	TM	1	
(0008,0034)	<i>Overlay Time</i>	TM	1	RET
(0008,0035)	<i>Curve Time</i>	TM	1	RET
(0008,0040)	<i>Data Set Type</i>	US	1	RET



(0008,0041)	Data Set Subtype	LO	1	RET
(0008,0042)	Nuclear Medicine Series Type	CS	1	RET
(0008,0050)	Accession Number	SH	1	
(0008,0052)	Query/Retrieve Level	CS	1	
(0008,0054)	Retrieve AE Title	AE	1-n	
(0008,0056)	Instance Availability	CS	1	
(0008,0058)	Failed SOP Instance UID List	UI	1-n	
(0008,0060)	Modality	CS	1	
(0008,0061)	Modalities in Study	CS	1-n	
(0008,0062)	SOP Classes in Study	UI	1-n	
(0008,0064)	Conversion Type	CS	1	
(0008,0068)	Presentation Intent Type	CS	1	
(0008,0070)	Manufacturer	LO	1	
(0008,0080)	Institution Name	LO	1	
(0008,0081)	Institution Address	ST	1	
(0008,0082)	Institution Code Sequence	SQ	1	
(0008,0090)	Referring Physician's Name	PN	1	
(0008,0092)	Referring Physician's Address	ST	1	
(0008,0094)	Referring Physician's Telephone Numbers	SH	1-n	
(0008,0096)	Referring Physician Identification Sequence	SQ	1	
(0008,0100)	Code Value	SH	1	
(0008,0102)	Coding Scheme Designator	SH	1	
(0008,0103)	Coding Scheme Version	SH	1	
(0008,0104)	Code Meaning	LO	1	
(0008,0105)	Mapping Resource	CS	1	
(0008,0106)	Context Group Version	DT	1	
(0008,0107)	Context Group Local Version	DT	1	
(0008,010B)	Context Group Extension Flag	CS	1	
(0008,010C)	Coding Scheme UID	UI	1	
(0008,010D)	Context Group Extension Creator UID	UI	1	
(0008,010F)	Context Identifier	CS	1	
(0008,0110)	Coding Scheme Identification Sequence	SQ	1	
(0008,0112)	Coding Scheme Registry	LO	1	
(0008,0114)	Coding Scheme External ID	ST	1	

(0008,0115)	Coding Scheme Name	ST	1	
(0008,0116)	Responsible Organization	ST	1	
(0008,0201)	Timezone Offset From UTC	SH	1	
<i>(0008,1000)</i>	<i>Network ID</i>	<i>AE</i>	<i>1</i>	<i>RET</i>
(0008,1010)	Station Name	SH	1	
(0008,1030)	Study Description	LO	1	
(0008,1032)	Procedure Code Sequence	SQ	1	
(0008,103E)	Series Description	LO	1	
(0008,1040)	Institutional Department Name	LO	1	
(0008,1048)	Physician(s) of Record	PN	1-n	
(0008,1049)	Physician(s) of Record Identification Sequence	SQ	1	
(0008,1050)	Performing Physician's Name	PN	1-n	
(0008,1052)	Performing Physician Identification Sequence	SQ	1	
(0008,1060)	Name of Physician(s) Reading Study	PN	1-n	
(0008,1062)	Physician(s) Reading Study Identification Sequence	SQ	1	
(0008,1070)	Operators' Name	PN	1-n	
(0008,1072)	Operator Identification Sequence	SQ	1	
(0008,1080)	Admitting Diagnoses Description	LO	1-n	
(0008,1084)	Admitting Diagnoses Code Sequence	SQ	1	
(0008,1090)	Manufacturer's Model Name	LO	1	
<i>(0008,1100)</i>	<i>Referenced Results Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
(0008,1110)	Referenced Study Sequence	SQ	1	
(0008,1111)	Referenced Performed Procedure Step Sequence	SQ	1	
(0008,1115)	Referenced Series Sequence	SQ	1	
(0008,1120)	Referenced Patient Sequence	SQ	1	
(0008,1125)	Referenced Visit Sequence	SQ	1	
<i>(0008,1130)</i>	<i>Referenced Overlay Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
(0008,113A)	Referenced Waveform Sequence	SQ	1	
(0008,1140)	Referenced Image Sequence	SQ	1	
<i>(0008,1145)</i>	<i>Referenced Curve Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
(0008,114A)	Referenced Instance Sequence	SQ	1	
(0008,114B)	Referenced Real World Value Mapping Instance Sequence	SQ	1	
(0008,1150)	Referenced SOP Class UID	UI	1	

(0008,1155)	Referenced SOP Instance UID	UI	1	
(0008,115A)	SOP Classes Supported	UI	1-n	
(0008,1160)	Referenced Frame Number	IS	1-n	
(0008,1195)	Transaction UID	UI	1	
(0008,1197)	Failure Reason	US	1	
(0008,1198)	Failed SOP Sequence	SQ	1	
(0008,1199)	Referenced SOP Sequence	SQ	1	
(0008,1200)	Studies Containing Other Referenced Instances Sequence	SQ	1	
(0008,1250)	Related Series Sequence	SQ	1	
(0008,2110)	<i>Lossy Image Compression</i>	CS	1	<i>RET</i>
(0008,2111)	Derivation Description	ST	1	
(0008,2112)	Source Image Sequence	SQ	1	
(0008,2120)	Stage Name	SH	1	
(0008,2122)	Stage Number	IS	1	
(0008,2124)	Number of Stages	IS	1	
(0008,2127)	View Name	SH	1	
(0008,2128)	View Number	IS	1	
(0008,2129)	Number of Event Timers	IS	1	
(0008,212A)	Number of Views in Stage	IS	1	
(0008,2130)	Event Elapsed Time(s)	DS	1-n	
(0008,2132)	Event Timer Name(s)	LO	1-n	
(0008,2142)	Start Trim	IS	1	
(0008,2143)	Stop Trim	IS	1	
(0008,2144)	Recommended Display Frame Rate	IS	1	
(0008,2200)	<i>Transducer Position</i>	CS	1	<i>RET</i>
(0008,2204)	<i>Transducer Orientation</i>	CS	1	<i>RET</i>
(0008,2208)	<i>Anatomic Structure</i>	CS	1	<i>RET</i>
(0008,2218)	Anatomic Region Sequence	SQ	1	
(0008,2220)	Anatomic Region Modifier Sequence	SQ	1	
(0008,2228)	Primary Anatomic Structure Sequence	SQ	1	
(0008,2229)	Anatomic Structure, Space or Region Sequence	SQ	1	
(0008,2230)	Primary Anatomic Structure Modifier Sequence	SQ	1	
(0008,2240)	<i>Transducer Position Sequence</i>	SQ	1	<i>RET</i>

(0008,2242)	Transducer Position Modifier Sequence	SQ	1	RET
(0008,2244)	Transducer Orientation Sequence	SQ	1	RET
(0008,2246)	Transducer Orientation Modifier Sequence	SQ	1	RET
(0008,3001)	Alternate Representation Sequence	SQ	1	
(0008,3010)	Irradiation Event UID	UI	1	
(0008,4000)	Identifying Comments	LT	1	RET
(0008,9007)	Frame Type	CS	4	
(0008,9092)	Referenced Image Evidence Sequence	SQ	1	
(0008,9121)	Referenced Raw Data Sequence	SQ	1	
(0008,9123)	Creator-Version UID	UI	1	
(0008,9124)	Derivation Image Sequence	SQ	1	
(0008,9154)	Source Image Evidence Sequence	SQ	1	
(0008,9205)	Pixel Presentation	CS	1	
(0008,9206)	Volumetric Properties	CS	1	
(0008,9207)	Volume Based Calculation Technique	CS	1	
(0008,9208)	Complex Image Component	CS	1	
(0008,9209)	Acquisition Contrast	CS	1	
(0008,9215)	Derivation Code Sequence	SQ	1	
(0008,9237)	Referenced Grayscale Presentation State Sequence	SQ	1	
(0008,9410)	Referenced Other Plane Sequence	SQ	1	
(0008,9458)	Frame Display Sequence	SQ	1	
(0008,9459)	Recommended Display Frame Rate in Float	FL	1	
(0008,9460)	Skip Frame Range Flag	CS	1	

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0010,0010)	Patient's Name	PN	1
(0010,0020)	Patient ID	LO	1
(0010,0021)	Issuer of Patient ID	LO	1
(0010,0022)	Type of Patient ID	CS	1
(0010,0030)	Patient's Birth Date	DA	1
(0010,0032)	Patient's Birth Time	TM	1
(0010,0040)	Patient's Sex	CS	1
(0010,0050)	Patient's Insurance Plan Code Sequence	SQ	1

(0010,0101)	Patient's Primary Language Code Sequence	SQ	1	
(0010,0102)	Patient's Primary Language Code Modifier Sequence	SQ	1	
(0010,1000)	Other Patient IDs	LO	1-n	
(0010,1001)	Other Patient Names	PN	1-n	
(0010,1002)	Other Patient IDs Sequence	SQ	1	
(0010,1005)	Patient's Birth Name	PN	1	
(0010,1010)	Patient's Age	AS	1	
(0010,1020)	Patient's Size	DS	1	
(0010,1030)	Patient's Weight	DS	1	
(0010,1040)	Patient's Address	LO	1	
(0010,1050)	<i>Insurance Plan Identification</i>	LO	1-n	RET
(0010,1060)	Patient's Mother's Birth Name	PN	1	
(0010,1080)	Military Rank	LO	1	
(0010,1081)	Branch of Service	LO	1	
(0010,1090)	Medical Record Locator	LO	1	
(0010,2000)	Medical Alerts	LO	1-n	
(0010,2110)	Contrast Allergies	LO	1-n	
(0010,2150)	Country of Residence	LO	1	
(0010,2152)	Region of Residence	LO	1	
(0010,2154)	Patient's Telephone Numbers	SH	1-n	
(0010,2160)	Ethnic Group	SH	1	
(0010,2180)	Occupation	SH	1	
(0010,21A0)	Smoking Status	CS	1	
(0010,21B0)	Additional Patient History	LT	1	
(0010,21C0)	Pregnancy Status	US	1	
(0010,21D0)	Last Menstrual Date	DA	1	
(0010,21F0)	Patient's Religious Preference	LO	1	
(0010,2201)	Patient Species Description	LO	1	
(0010,2202)	Patient Species Code Sequence	SQ	1	
(0010,2203)	Patient's Sex Neutered	CS	1	
(0010,2292)	Patient Breed Description	LO	1	
(0010,2293)	Patient Breed Code Sequence	SQ	1	
(0010,2294)	Breed Registration Sequence	SQ	1	
(0010,2295)	Breed Registration Number	LO	1	

(0010,2296)	Breed Registry Code Sequence	SQ	1
(0010,2297)	Responsible Person	PN	1
(0010,2298)	Responsible Person Role	CS	1
(0010,2299)	Responsible Organization	LO	1
(0010,4000)	Patient Comments	LT	1
(0010,9431)	Examined Body Thickness	FL	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0012,0010)	Clinical Trial Sponsor Name	LO	1
(0012,0020)	Clinical Trial Protocol ID	LO	1
(0012,0021)	Clinical Trial Protocol Name	LO	1
(0012,0030)	Clinical Trial Site ID	LO	1
(0012,0031)	Clinical Trial Site Name	LO	1
(0012,0040)	Clinical Trial Subject ID	LO	1
(0012,0042)	Clinical Trial Subject Reading ID	LO	1
(0012,0050)	Clinical Trial Time Point ID	LO	1
(0012,0051)	Clinical Trial Time Point Description	ST	1
(0012,0060)	Clinical Trial Coordinating Center Name	LO	1
(0012,0062)	Patient Identity Removed	CS	1
(0012,0063)	De-identification Method	LO	1-n
(0012,0064)	De-identification Method Code Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0018,0010)	Contrast/Bolus Agent	LO	1
(0018,0012)	Contrast/Bolus Agent Sequence	SQ	1
(0018,0014)	Contrast/Bolus Administration Route Sequence	SQ	1
(0018,0015)	Body Part Examined	CS	1
(0018,0020)	Scanning Sequence	CS	1-n
(0018,0021)	Sequence Variant	CS	1-n
(0018,0022)	Scan Options	CS	1-n
(0018,0023)	MR Acquisition Type	CS	1
(0018,0024)	Sequence Name	SH	1
(0018,0025)	Angio Flag	CS	1

(0018,0026)	Intervention Drug Information Sequence	SQ	1	
(0018,0027)	Intervention Drug Stop Time	TM	1	
(0018,0028)	Intervention Drug Dose	DS	1	
(0018,0029)	Intervention Drug Sequence	SQ	1	
(0018,002A)	Additional Drug Sequence	SQ	1	
(0018,0030)	<i>Radionuclide</i>	LO	1-n	RET
(0018,0031)	Radiopharmaceutical	LO	1	
(0018,0032)	<i>Energy Window Centerline</i>	DS	1	RET
(0018,0033)	<i>Energy Window Total Width</i>	DS	1-n	RET
(0018,0034)	Intervention Drug Name	LO	1	
(0018,0035)	Intervention Drug Start Time	TM	1	
(0018,0036)	Intervention Sequence	SQ	1	
(0018,0037)	<i>Therapy Type</i>	CS	1	RET
(0018,0038)	Intervention Status	CS	1	
(0018,0039)	<i>Therapy Description</i>	CS	1	RET
(0018,003A)	Intervention Description	ST	1	
(0018,0040)	Cine Rate	IS	1	
(0018,0050)	Slice Thickness	DS	1	
(0018,0060)	KVP	DS	1	
(0018,0070)	Counts Accumulated	IS	1	
(0018,0071)	Acquisition Termination Condition	CS	1	
(0018,0072)	Effective Duration	DS	1	
(0018,0073)	Acquisition Start Condition	CS	1	
(0018,0074)	Acquisition Start Condition Data	IS	1	
(0018,0075)	Acquisition Termination Condition Data	IS	1	
(0018,0080)	Repetition Time	DS	1	
(0018,0081)	Echo Time	DS	1	
(0018,0082)	Inversion Time	DS	1	
(0018,0083)	Number of Averages	DS	1	
(0018,0084)	Imaging Frequency	DS	1	
(0018,0085)	Imaged Nucleus	SH	1	
(0018,0086)	Echo Number(s)	IS	1-n	
(0018,0087)	Magnetic Field Strength	DS	1	
(0018,0088)	Spacing Between Slices	DS	1	

(0018,0089)	Number of Phase Encoding Steps	IS	1
(0018,0090)	Data Collection Diameter	DS	1
(0018,0091)	Echo Train Length	IS	1
(0018,0093)	Percent Sampling	DS	1
(0018,0094)	Percent Phase Field of View	DS	1
(0018,0095)	Pixel Bandwidth	DS	1
(0018,1000)	Device Serial Number	LO	1
(0018,1002)	Device UID	UI	1
(0018,1003)	Device ID	LO	1
(0018,1004)	Plate ID	LO	1
(0018,1005)	Generator ID	LO	1
(0018,1006)	Grid ID	LO	1
(0018,1007)	Cassette ID	LO	1
(0018,1008)	Gantry ID	LO	1
(0018,1010)	Secondary Capture Device ID	LO	1
(0018,1011)	Hardcopy Creation Device ID	LO	1
(0018,1012)	Date of Secondary Capture	DA	1
(0018,1014)	Time of Secondary Capture	TM	1
(0018,1016)	Secondary Capture Device Manufacturer	LO	1
(0018,1017)	Hardcopy Device Manufacturer	LO	1
(0018,1018)	Secondary Capture Device Manufacturer's Model Name	LO	1
(0018,1019)	Secondary Capture Device Software Version(s)	LO	1-n
(0018,101A)	Hardcopy Device Software Version	LO	1-n
(0018,101B)	Hardcopy Device Manufacturer's Model Name	LO	1
(0018,1020)	Software Version(s)	LO	1-n
(0018,1022)	Video Image Format Acquired	SH	1
(0018,1023)	Digital Image Format Acquired	LO	1
(0018,1030)	Protocol Name	LO	1
(0018,1040)	Contrast/Bolus Route	LO	1
(0018,1041)	Contrast/Bolus Volume	DS	1
(0018,1042)	Contrast/Bolus Start Time	TM	1
(0018,1043)	Contrast/Bolus Stop Time	TM	1
(0018,1044)	Contrast/Bolus Total Dose	DS	1
(0018,1045)	Syringe Counts	IS	1



(0018,1046)	Contrast Flow Rate	DS	1-n
(0018,1047)	Contrast Flow Duration	DS	1-n
(0018,1048)	Contrast/Bolus Ingredient	CS	1
(0018,1049)	Contrast/Bolus Ingredient Concentration	DS	1
(0018,1050)	Spatial Resolution	DS	1
(0018,1060)	Trigger Time	DS	1
(0018,1061)	Trigger Source or Type	LO	1
(0018,1062)	Nominal Interval	IS	1
(0018,1063)	Frame Time	DS	1
(0018,1064)	Framing Type	LO	1
(0018,1065)	Frame Time Vector	DS	1-n
(0018,1066)	Frame Delay	DS	1
(0018,1067)	Image Trigger Delay	DS	1
(0018,1068)	Multiplex Group Time Offset	DS	1
(0018,1069)	Trigger Time Offset	DS	1
(0018,106A)	Synchronization Trigger	CS	1
(0018,106C)	Synchronization Channel	US	2
(0018,106E)	Trigger Sample Position	UL	1
(0018,1070)	Radiopharmaceutical Route	LO	1
(0018,1071)	Radiopharmaceutical Volume	DS	1
(0018,1072)	Radiopharmaceutical Start Time	TM	1
(0018,1073)	Radiopharmaceutical Stop Time	TM	1
(0018,1074)	Radionuclide Total Dose	DS	1
(0018,1075)	Radionuclide Half Life	DS	1
(0018,1076)	Radionuclide Positron Fraction	DS	1
(0018,1077)	Radiopharmaceutical Specific Activity	DS	1
(0018,1078)	Radiopharmaceutical Start Datetime	DT	1
(0018,1079)	Radiopharmaceutical Stop Datetime	DT	1
(0018,1080)	Beat Rejection Flag	CS	1
(0018,1081)	Low R-R Value	IS	1
(0018,1082)	High R-R Value	IS	1
(0018,1083)	Intervals Acquired	IS	1
(0018,1084)	Intervals Rejected	IS	1
(0018,1085)	PVC Rejection	LO	1

(0018,1086)	Skip Beats	IS	1	
(0018,1088)	Heart Rate	IS	1	
(0018,1090)	Cardiac Number of Images	IS	1	
(0018,1094)	Trigger Window	IS	1	
(0018,1100)	Reconstruction Diameter	DS	1	
(0018,1110)	Distance Source to Detector	DS	1	
(0018,1111)	Distance Source to Patient	DS	1	
(0018,1114)	Estimated Radiographic Magnification Factor	DS	1	
(0018,1120)	Gantry/Detector Tilt	DS	1	
(0018,1121)	Gantry/Detector Slew	DS	1	
(0018,1130)	Table Height	DS	1	
(0018,1131)	Table Traverse	DS	1	
(0018,1134)	Table Motion	CS	1	
(0018,1135)	Table Vertical Increment	DS	1-n	
(0018,1136)	Table Lateral Increment	DS	1-n	
(0018,1137)	Table Longitudinal Increment	DS	1-n	
(0018,1138)	Table Angle	DS	1	
(0018,113A)	Table Type	CS	1	
(0018,1140)	Rotation Direction	CS	1	
(0018,1141)	Angular Position	DS	1	
(0018,1142)	Radial Position	DS	1-n	
(0018,1143)	Scan Arc	DS	1	
(0018,1144)	Angular Step	DS	1	
(0018,1145)	Center of Rotation Offset	DS	1	
(0018,1146)	<i>Rotation Offset</i>	<i>DS</i>	<i>1-n</i>	<i>RET</i>
(0018,1147)	Field of View Shape	CS	1	
(0018,1149)	Field of View Dimension(s)	IS	1-2	
(0018,1150)	Exposure Time	IS	1	
(0018,1151)	X-ray Tube Current	IS	1	
(0018,1152)	Exposure	IS	1	
(0018,1153)	Exposure in uAs	IS	1	
(0018,1154)	Average Pulse Width	DS	1	
(0018,1155)	Radiation Setting	CS	1	
(0018,1156)	Rectification Type	CS	1	

(0018,115A)	Radiation Mode	CS	1	
(0018,115E)	Image and Fluoroscopy Area Dose Product	DS	1	
(0018,1160)	Filter Type	SH	1	
(0018,1161)	Type of Filters	LO	1-n	
(0018,1162)	Intensifier Size	DS	1	
(0018,1164)	Imager Pixel Spacing	DS	2	
(0018,1166)	Grid	CS	1-n	
(0018,1170)	Generator Power	IS	1	
(0018,1180)	Collimator/grid Name	SH	1	
(0018,1181)	Collimator Type	CS	1	
(0018,1182)	Focal Distance	IS	1-2	
(0018,1183)	X Focus Center	DS	1-2	
(0018,1184)	Y Focus Center	DS	1-2	
(0018,1190)	Focal Spot(s)	DS	1-n	
(0018,1191)	Anode Target Material	CS	1	
(0018,11A0)	Body Part Thickness	DS	1	
(0018,11A2)	Compression Force	DS	1	
(0018,1200)	Date of Last Calibration	DA	1-n	
(0018,1201)	Time of Last Calibration	TM	1-n	
(0018,1210)	Convolution Kernel	SH	1-n	
(0018,1240)	<i>Upper/Lower Pixel Values</i>	<i>IS</i>	<i>1-n</i>	<i>RET</i>
(0018,1242)	Actual Frame Duration	IS	1	
(0018,1243)	Count Rate	IS	1	
(0018,1244)	Preferred Playback Sequencing	US	1	
(0018,1250)	Receive Coil Name	SH	1	
(0018,1251)	Transmit Coil Name	SH	1	
(0018,1260)	Plate Type	SH	1	
(0018,1261)	Phosphor Type	LO	1	
(0018,1300)	Scan Velocity	DS	1	
(0018,1301)	Whole Body Technique	CS	1-n	
(0018,1302)	Scan Length	IS	1	
(0018,1310)	Acquisition Matrix	US	4	
(0018,1312)	In-plane Phase Encoding Direction	CS	1	
(0018,1314)	Flip Angle	DS	1	

(0018,1315)	Variable Flip Angle Flag	CS	1
(0018,1316)	SAR	DS	1
(0018,1318)	dB/dt	DS	1
(0018,1400)	Acquisition Device Processing Description	LO	1
(0018,1401)	Acquisition Device Processing Code	LO	1
(0018,1402)	Cassette Orientation	CS	1
(0018,1403)	Cassette Size	CS	1
(0018,1404)	Exposures on Plate	US	1
(0018,1405)	Relative X-ray Exposure	IS	1
(0018,1450)	Column Angulation	DS	1
(0018,1460)	Tomo Layer Height	DS	1
(0018,1470)	Tomo Angle	DS	1
(0018,1480)	Tomo Time	DS	1
(0018,1490)	Tomo Type	CS	1
(0018,1491)	Tomo Class	CS	1
(0018,1495)	Number of Tomosynthesis Source Images	IS	1
(0018,1500)	Positioner Motion	CS	1
(0018,1508)	Positioner Type	CS	1
(0018,1510)	Positioner Primary Angle	DS	1
(0018,1511)	Positioner Secondary Angle	DS	1
(0018,1520)	Positioner Primary Angle Increment	DS	1-n
(0018,1521)	Positioner Secondary Angle Increment	DS	1-n
(0018,1530)	Detector Primary Angle	DS	1
(0018,1531)	Detector Secondary Angle	DS	1
(0018,1600)	Shutter Shape	CS	1-3
(0018,1602)	Shutter Left Vertical Edge	IS	1
(0018,1604)	Shutter Right Vertical Edge	IS	1
(0018,1606)	Shutter Upper Horizontal Edge	IS	1
(0018,1608)	Shutter Lower Horizontal Edge	IS	1
(0018,1610)	Center of Circular Shutter	IS	2
(0018,1612)	Radius of Circular Shutter	IS	1
(0018,1620)	Vertices of the Polygonal Shutter	IS	2-2n
(0018,1622)	Shutter Presentation Value	US	1
(0018,1623)	Shutter Overlay Group	US	1

(0018,1624)	Shutter Presentation Color CIELab Value	US	3	
(0018,1700)	Collimator Shape	CS	1-3	
(0018,1702)	Collimator Left Vertical Edge	IS	1	
(0018,1704)	Collimator Right Vertical Edge	IS	1	
(0018,1706)	Collimator Upper Horizontal Edge	IS	1	
(0018,1708)	Collimator Lower Horizontal Edge	IS	1	
(0018,1710)	Center of Circular Collimator	IS	2	
(0018,1712)	Radius of Circular Collimator	IS	1	
(0018,1720)	Vertices of the Polygonal Collimator	IS	2-2n	
(0018,1800)	Acquisition Time Synchronized	CS	1	
(0018,1801)	Time Source	SH	1	
(0018,1802)	Time Distribution Protocol	CS	1	
(0018,1803)	NTP Source Address	LO	1	
(0018,2001)	Page Number Vector	IS	1-n	
(0018,2002)	Frame Label Vector	SH	1-n	
(0018,2003)	Frame Primary Angle Vector	DS	1-n	
(0018,2004)	Frame Secondary Angle Vector	DS	1-n	
(0018,2005)	Slice Location Vector	DS	1-n	
(0018,2006)	Display Window Label Vector	SH	1-n	
(0018,2010)	Nominal Scanned Pixel Spacing	DS	2	
(0018,2020)	Digitizing Device Transport Direction	CS	1	
(0018,2030)	Rotation of Scanned Film	DS	1	
(0018,3100)	IVUS Acquisition	CS	1	
(0018,3101)	IVUS Pullback Rate	DS	1	
(0018,3102)	IVUS Gated Rate	DS	1	
(0018,3103)	IVUS Pullback Start Frame Number	IS	1	
(0018,3104)	IVUS Pullback Stop Frame Number	IS	1	
(0018,3105)	Lesion Number	IS	1-n	
(0018,4000)	<i>Acquisition Comments</i>	<i>LT</i>	<i>1</i>	<i>RET</i>
(0018,5000)	Output Power	SH	1-n	
(0018,5010)	Transducer Data	LO	3	
(0018,5012)	Focus Depth	DS	1	
(0018,5020)	Processing Function	LO	1	
(0018,5021)	Postprocessing Function	LO	1	

(0018,5022)	Mechanical Index	DS	1	
(0018,5024)	Bone Thermal Index	DS	1	
(0018,5026)	Cranial Thermal Index	DS	1	
(0018,5027)	Soft Tissue Thermal Index	DS	1	
(0018,5028)	Soft Tissue-focus Thermal Index	DS	1	
(0018,5029)	Soft Tissue-surface Thermal Index	DS	1	
(0018,5030)	<i>Dynamic Range</i>	<i>DS</i>	<i>1</i>	<i>RET</i>
(0018,5040)	<i>Total Gain</i>	<i>DS</i>	<i>1</i>	<i>RET</i>
(0018,5050)	Depth of Scan Field	IS	1	
(0018,5100)	Patient Position	CS	1	
(0018,5101)	View Position	CS	1	
(0018,5104)	Projection Eponymous Name Code Sequence	SQ	1	
(0018,5210)	<i>Image Transformation Matrix</i>	<i>DS</i>	<i>6</i>	<i>RET</i>
(0018,5212)	<i>Image Translation Vector</i>	<i>DS</i>	<i>3</i>	<i>RET</i>
(0018,6000)	Sensitivity	DS	1	
(0018,6011)	Sequence of Ultrasound Regions	SQ	1	
(0018,6012)	Region Spatial Format	US	1	
(0018,6014)	Region Data Type	US	1	
(0018,6016)	Region Flags	UL	1	
(0018,6018)	Region Location Min $X_0$	UL	1	
(0018,601A)	Region Location Min $Y_0$	UL	1	
(0018,601C)	Region Location Max $X_1$	UL	1	
(0018,601E)	Region Location Max $Y_1$	UL	1	
(0018,6020)	Reference Pixel $X_0$	SL	1	
(0018,6022)	Reference Pixel $Y_0$	SL	1	
(0018,6024)	Physical Units X Direction	US	1	
(0018,6026)	Physical Units Y Direction	US	1	
(0018,6028)	Reference Pixel Physical Value X	FD	1	
(0018,602A)	Reference Pixel Physical Value Y	FD	1	
(0018,602C)	Physical Delta X	FD	1	
(0018,602E)	Physical Delta Y	FD	1	
(0018,6030)	Transducer Frequency	UL	1	
(0018,6031)	Transducer Type	CS	1	

(0018,6032)	Pulse Repetition Frequency	UL	1	
(0018,6034)	Doppler Correction Angle	FD	1	
(0018,6036)	Steering Angle	FD	1	
(0018,6038)	<i>Doppler Sample Volume X Position</i>	UL	1	RET
(0018,6039)	Doppler Sample Volume X Position	SL	1	
(0018,603A)	<i>Doppler Sample Volume Y Position</i>	UL	1	RET
(0018,603B)	Doppler Sample Volume Y Position	SL	1	
(0018,603C)	<i>TM-Line Position X<sub>0</sub></i>	UL	1	RET
(0018,603D)	TM-Line Position X <sub>0</sub>	SL	1	
(0018,603E)	<i>TM-Line Position Y<sub>0</sub></i>	UL	1	RET
(0018,603F)	TM-Line Position Y <sub>0</sub>	SL	1	
(0018,6040)	<i>TM-Line Position X<sub>1</sub></i>	UL	1	RET
(0018,6041)	TM-Line Position X <sub>1</sub>	SL	1	
(0018,6042)	<i>TM-Line Position Y<sub>1</sub></i>	UL	1	RET
(0018,6043)	TM-Line Position Y <sub>1</sub>	SL	1	
(0018,6044)	Pixel Component Organization	US	1	
(0018,6046)	Pixel Component Mask	UL	1	
(0018,6048)	Pixel Component Range Start	UL	1	
(0018,604A)	Pixel Component Range Stop	UL	1	
(0018,604C)	Pixel Component Physical Units	US	1	
(0018,604E)	Pixel Component Data Type	US	1	
(0018,6050)	Number of Table Break Points	UL	1	
(0018,6052)	Table of X Break Points	UL	1-n	
(0018,6054)	Table of Y Break Points	FD	1-n	
(0018,6056)	Number of Table Entries	UL	1	
(0018,6058)	Table of Pixel Values	UL	1-n	
(0018,605A)	Table of Parameter Values	FL	1-n	
(0018,6060)	R Wave Time Vector	FL	1-n	
(0018,7000)	Detector Conditions Nominal Flag	CS	1	
(0018,7001)	Detector Temperature	DS	1	
(0018,7004)	Detector Type	CS	1	
(0018,7005)	Detector Configuration	CS	1	
(0018,7006)	Detector Description	LT	1	

(0018,7008)	Detector Mode	LT	1
(0018,700A)	Detector ID	SH	1
(0018,700C)	Date of Last Detector Calibration	DA	1
(0018,700E)	Time of Last Detector Calibration	TM	1
(0018,7010)	Exposures on Detector Since Last Calibration	IS	1
(0018,7011)	Exposures on Detector Since Manufactured	IS	1
(0018,7012)	Detector Time Since Last Exposure	DS	1
(0018,7014)	Detector Active Time	DS	1
(0018,7016)	Detector Activation Offset From Exposure	DS	1
(0018,701A)	Detector Binning	DS	2
(0018,7020)	Detector Element Physical Size	DS	2
(0018,7022)	Detector Element Spacing	DS	2
(0018,7024)	Detector Active Shape	CS	1
(0018,7026)	Detector Active Dimension(s)	DS	1-2
(0018,7028)	Detector Active Origin	DS	2
(0018,702A)	Detector Manufacturer Name	LO	1
(0018,702B)	Detector Manufacturer's Model Name	LO	1
(0018,7030)	Field of View Origin	DS	2
(0018,7032)	Field of View Rotation	DS	1
(0018,7034)	Field of View Horizontal Flip	CS	1
(0018,7040)	Grid Absorbing Material	LT	1
(0018,7041)	Grid Spacing Material	LT	1
(0018,7042)	Grid Thickness	DS	1
(0018,7044)	Grid Pitch	DS	1
(0018,7046)	Grid Aspect Ratio	IS	2
(0018,7048)	Grid Period	DS	1
(0018,704C)	Grid Focal Distance	DS	1
(0018,7050)	Filter Material	CS	1-n
(0018,7052)	Filter Thickness Minimum	DS	1-n
(0018,7054)	Filter Thickness Maximum	DS	1-n
(0018,7060)	Exposure Control Mode	CS	1
(0018,7062)	Exposure Control Mode Description	LT	1
(0018,7064)	Exposure Status	CS	1
(0018,7065)	Phototimer Setting	DS	1



(0018,8150)	Exposure Time in $\mu$ S	DS	1
(0018,8151)	X-Ray Tube Current in $\mu$ A	DS	1
(0018,9004)	Content Qualification	CS	1
(0018,9005)	Pulse Sequence Name	SH	1
(0018,9006)	MR Imaging Modifier Sequence	SQ	1
(0018,9008)	Echo Pulse Sequence	CS	1
(0018,9009)	Inversion Recovery	CS	1
(0018,9010)	Flow Compensation	CS	1
(0018,9011)	Multiple Spin Echo	CS	1
(0018,9012)	Multi-planar Excitation	CS	1
(0018,9014)	Phase Contrast	CS	1
(0018,9015)	Time of Flight Contrast	CS	1
(0018,9016)	Spoiling	CS	1
(0018,9017)	Steady State Pulse Sequence	CS	1
(0018,9018)	Echo Planar Pulse Sequence	CS	1
(0018,9019)	Tag Angle First Axis	FD	1
(0018,9020)	Magnetization Transfer	CS	1
(0018,9021)	T2 Preparation	CS	1
(0018,9022)	Blood Signal Nulling	CS	1
(0018,9024)	Saturation Recovery	CS	1
(0018,9025)	Spectrally Selected Suppression	CS	1
(0018,9026)	Spectrally Selected Excitation	CS	1
(0018,9027)	Spatial Pre-saturation	CS	1
(0018,9028)	Tagging	CS	1
(0018,9029)	Oversampling Phase	CS	1
(0018,9030)	Tag Spacing First Dimension	FD	1
(0018,9032)	Geometry of k-Space Traversal	CS	1
(0018,9033)	Segmented k-Space Traversal	CS	1
(0018,9034)	Rectilinear Phase Encode Reordering	CS	1
(0018,9035)	Tag Thickness	FD	1
(0018,9036)	Partial Fourier Direction	CS	1
(0018,9037)	Cardiac Synchronization Technique	CS	1
(0018,9041)	Receive Coil Manufacturer Name	LO	1
(0018,9042)	MR Receive Coil Sequence	SQ	1

(0018,9043)	Receive Coil Type	CS	1
(0018,9044)	Quadrature Receive Coil	CS	1
(0018,9045)	Multi-Coil Definition Sequence	SQ	1
(0018,9046)	Multi-Coil Configuration	LO	1
(0018,9047)	Multi-Coil Element Name	SH	1
(0018,9048)	Multi-Coil Element Used	CS	1
(0018,9049)	MR Transmit Coil Sequence	SQ	1
(0018,9050)	Transmit Coil Manufacturer Name	LO	1
(0018,9051)	Transmit Coil Type	CS	1
(0018,9052)	Spectral Width	FD	1-2
(0018,9053)	Chemical Shift Reference	FD	1-2
(0018,9054)	Volume Localization Technique	CS	1
(0018,9058)	MR Acquisition Frequency Encoding Steps	US	1
(0018,9059)	De-coupling	CS	1
(0018,9060)	De-coupled Nucleus	CS	1-2
(0018,9061)	De-coupling Frequency	FD	1-2
(0018,9062)	De-coupling Method	CS	1
(0018,9063)	De-coupling Chemical Shift Reference	FD	1-2
(0018,9064)	k-space Filtering	CS	1
(0018,9065)	Time Domain Filtering	CS	1-2
(0018,9066)	Number of Zero fills	US	1-2
(0018,9067)	Baseline Correction	CS	1
(0018,9069)	Parallel Reduction Factor In-plane	FD	1
(0018,9070)	Cardiac R-R Interval Specified	FD	1
(0018,9073)	Acquisition Duration	FD	1
(0018,9074)	Frame Acquisition Datetime	DT	1
(0018,9075)	Diffusion Directionality	CS	1
(0018,9076)	Diffusion Gradient Direction Sequence	SQ	1
(0018,9077)	Parallel Acquisition	CS	1
(0018,9078)	Parallel Acquisition Technique	CS	1
(0018,9079)	Inversion Times	FD	1-n
(0018,9080)	Metabolite Map Description	ST	1
(0018,9081)	Partial Fourier	CS	1
(0018,9082)	Effective Echo Time	FD	1

(0018,9083)	Metabolite Map Code Sequence	SQ	1
(0018,9084)	Chemical Shift Sequence	SQ	1
(0018,9085)	Cardiac Signal Source	CS	1
(0018,9087)	Diffusion b-value	FD	1
(0018,9089)	Diffusion Gradient Orientation	FD	3
(0018,9090)	Velocity Encoding Direction	FD	3
(0018,9091)	Velocity Encoding Minimum Value	FD	1
(0018,9093)	Number of k-Space Trajectories	US	1
(0018,9094)	Coverage of k-Space	CS	1
(0018,9095)	Spectroscopy Acquisition Phase Rows	UL	1
(0018,9098)	Transmitter Frequency	FD	1-2
(0018,9100)	Resonant Nucleus	CS	1-2
(0018,9101)	Frequency Correction	CS	1
(0018,9103)	MR Spectroscopy FOV/Geometry Sequence	SQ	1
(0018,9104)	Slab Thickness	FD	1
(0018,9105)	Slab Orientation	FD	3
(0018,9106)	Mid Slab Position	FD	3
(0018,9107)	MR Spatial Saturation Sequence	SQ	1
(0018,9112)	MR Timing and Related Parameters Sequence	SQ	1
(0018,9114)	MR Echo Sequence	SQ	1
(0018,9115)	MR Modifier Sequence	SQ	1
(0018,9117)	MR Diffusion Sequence	SQ	1
(0018,9118)	Cardiac Trigger Sequence	SQ	1
(0018,9119)	MR Averages Sequence	SQ	1
(0018,9125)	MR FOV/Geometry Sequence	SQ	1
(0018,9126)	Volume Localization Sequence	SQ	1
(0018,9127)	Spectroscopy Acquisition Data Columns	UL	1
(0018,9147)	Diffusion Anisotropy Type	CS	1
(0018,9151)	Frame Reference Datetime	DT	1
(0018,9152)	MR Metabolite Map Sequence	SQ	1
(0018,9155)	Parallel Reduction Factor out-of-plane	FD	1
(0018,9159)	Spectroscopy Acquisition Out-of-plane Phase Steps	UL	1
(0018,9166)	Bulk Motion Status	CS	1
(0018,9168)	Parallel Reduction Factor Second In-plane	FD	1

(0018,9169)	Cardiac Beat Rejection Technique	CS	1	
(0018,9170)	Respiratory Motion Compensation Technique	CS	1	
(0018,9171)	Respiratory Signal Source	CS	1	
(0018,9172)	Bulk Motion Compensation Technique	CS	1	
(0018,9173)	Bulk Motion Signal Source	CS	1	
(0018,9174)	Applicable Safety Standard Agency	CS	1	
(0018,9175)	Applicable Safety Standard Description	LO	1	
(0018,9176)	Operating Mode Sequence	SQ	1	
(0018,9177)	Operating Mode Type	CS	1	
(0018,9178)	Operating Mode	CS	1	
(0018,9179)	Specific Absorption Rate Definition	CS	1	
(0018,9180)	Gradient Output Type	CS	1	
(0018,9181)	Specific Absorption Rate Value	FD	1	
(0018,9182)	Gradient Output	FD	1	
(0018,9183)	Flow Compensation Direction	CS	1	
(0018,9184)	Tagging Delay	FD	1	
(0018,9185)	Respiratory Motion Compensation Technique Description	ST	1	
(0018,9186)	Respiratory Signal Source ID	SH	1	
(0018,9195)	<i>Chemical Shifts Minimum Integration Limit in Hz</i>	<i>FD</i>	<i>1</i>	<i>RET</i>
(0018,9196)	<i>Chemical Shifts Maximum Integration Limit in Hz</i>	<i>FD</i>	<i>1</i>	<i>RET</i>
(0018,9197)	MR Velocity Encoding Sequence	SQ	1	
(0018,9198)	First Order Phase Correction	CS	1	
(0018,9199)	Water Referenced Phase Correction	CS	1	
(0018,9200)	MR Spectroscopy Acquisition Type	CS	1	
(0018,9214)	Respiratory Cycle Position	CS	1	
(0018,9217)	Velocity Encoding Maximum Value	FD	1	
(0018,9218)	Tag Spacing Second Dimension	FD	1	
(0018,9219)	Tag Angle Second Axis	SS	1	
(0018,9220)	Frame Acquisition Duration	FD	1	
(0018,9226)	MR Image Frame Type Sequence	SQ	1	
(0018,9227)	MR Spectroscopy Frame Type Sequence	SQ	1	
(0018,9231)	MR Acquisition Phase Encoding Steps in-plane	US	1	
(0018,9232)	MR Acquisition Phase Encoding Steps out-of-plane	US	1	

(0018,9234)	Spectroscopy Acquisition Phase Columns	UL	1
(0018,9236)	Cardiac Cycle Position	CS	1
(0018,9239)	Specific Absorption Rate Sequence	SQ	1
(0018,9240)	RF Echo Train Length	US	1
(0018,9241)	Gradient Echo Train Length	US	1
(0018,9295)	Chemical Shifts Minimum Integration Limit in ppm	FD	1
(0018,9296)	Chemical Shifts Maximum Integration Limit in ppm	FD	1
(0018,9301)	CT Acquisition Type Sequence	SQ	1
(0018,9302)	Acquisition Type	CS	1
(0018,9303)	Tube Angle	FD	1
(0018,9304)	CT Acquisition Details Sequence	SQ	1
(0018,9305)	Revolution Time	FD	1
(0018,9306)	Single Collimation Width	FD	1
(0018,9307)	Total Collimation Width	FD	1
(0018,9308)	CT Table Dynamics Sequence	SQ	1
(0018,9309)	Table Speed	FD	1
(0018,9310)	Table Feed per Rotation	FD	1
(0018,9311)	Spiral Pitch Factor	FD	1
(0018,9312)	CT Geometry Sequence	SQ	1
(0018,9313)	Data Collection Center (Patient)	FD	3
(0018,9314)	CT Reconstruction Sequence	SQ	1
(0018,9315)	Reconstruction Algorithm	CS	1
(0018,9316)	Convolution Kernel Group	CS	1
(0018,9317)	Reconstruction Field of View	FD	2
(0018,9318)	Reconstruction Target Center (Patient)	FD	3
(0018,9319)	Reconstruction Angle	FD	1
(0018,9320)	Image Filter	SH	1
(0018,9321)	CT Exposure Sequence	SQ	1
(0018,9322)	Reconstruction Pixel Spacing	FD	2
(0018,9323)	Exposure Modulation Type	CS	1
(0018,9324)	Estimated Dose Saving	FD	1
(0018,9325)	CT X-ray Details Sequence	SQ	1
(0018,9326)	CT Position Sequence	SQ	1
(0018,9327)	Table Position	FD	1

(0018,9328)	Exposure Time in ms	FD	1
(0018,9329)	CT Image Frame Type Sequence	SQ	1
(0018,9330)	X-Ray Tube Current in mA	FD	1
(0018,9332)	Exposure in mAs	FD	1
(0018,9333)	Constant Volume Flag	CS	1
(0018,9334)	Fluoroscopy Flag	CS	1
(0018,9335)	Distance Source to Data Collection Center	FD	1
(0018,9337)	Contrast/Bolus Agent Number	US	1
(0018,9338)	Contrast/Bolus Ingredient Code Sequence	SQ	1
(0018,9340)	Contrast Administration Profile Sequence	SQ	1
(0018,9341)	Contrast/Bolus Usage Sequence	SQ	1
(0018,9342)	Contrast/Bolus Agent Administered	CS	1
(0018,9343)	Contrast/Bolus Agent Detected	CS	1
(0018,9344)	Contrast/Bolus Agent Phase	CS	1
(0018,9345)	CTDIvol	FD	1
(0018,9401)	Projection Pixel Calibration Sequence	SQ	1
(0018,9402)	Distance Source to Isocenter	FL	1
(0018,9403)	Distance Object to Table Top	FL	1
(0018,9404)	Object Pixel Spacing in Center of Beam	FL	2
(0018,9405)	Positioner Position Sequence	SQ	1
(0018,9406)	Table Position Sequence	SQ	1
(0018,9407)	Collimator Shape Sequence	SQ	1
(0018,9412)	XA/XRF Frame Characteristics Sequence	SQ	1
(0018,9417)	Frame Acquisition Sequence	SQ	1
(0018,9420)	X-Ray Receptor Type	CS	1
(0018,9423)	Acquisition Protocol Name	LO	1
(0018,9424)	Acquisition Protocol Description	LT	1
(0018,9425)	Contrast/Bolus Ingredient Opaque	CS	1
(0018,9426)	Distance Receptor Plane to Detector Housing	FL	1
(0018,9427)	Intensifier Active Shape	CS	1
(0018,9428)	Intensifier Active Dimension(s)	FL	1-2
(0018,9429)	Physical Detector Size	FL	2
(0018,9430)	Position of Isocenter Projection	US	2
(0018,9432)	Field of View Sequence	SQ	1

(0018,9433)	Field of View Description	LO	1
(0018,9434)	Exposure Control Sensing Regions Sequence	SQ	1
(0018,9435)	Exposure Control Sensing Region Shape	CS	1
(0018,9436)	Exposure Control Sensing Region Left Vertical Edge	SS	1
(0018,9437)	Exposure Control Sensing Region Right Vertical Edge	SS	1
(0018,9438)	Exposure Control Sensing Region Upper Horizontal Edge	SS	1
(0018,9439)	Exposure Control Sensing Region Lower Horizontal Edge	SS	1
(0018,9440)	Center of Circular Exposure Control Sensing Region	SS	2
(0018,9441)	Radius of Circular Exposure Control Sensing Region	US	1
(0018,9442)	Vertices of the Polygonal Exposure Control Sensing Region	SS	2-n
(0018,9445)			<i>RET</i>
(0018,9447)	Column Angulation (Patient)	FL	1
(0018,9449)	Beam Angle	FL	1
(0018,9451)	Frame Detector Parameters Sequence	SQ	1
(0018,9452)	Calculated Anatomy Thickness	FL	1
(0018,9455)	Calibration Sequence	SQ	1
(0018,9456)	Object Thickness Sequence	SQ	1
(0018,9457)	Plane Identification	CS	1
(0018,9461)	Field of View Dimension(s) in Float	FL	1-2
(0018,9462)	Isocenter Reference System Sequence	SQ	1
(0018,9463)	Positioner Isocenter Primary Angle	FL	1
(0018,9464)	Positioner Isocenter Secondary Angle	FL	1
(0018,9465)	Positioner Isocenter Detector Rotation Angle	FL	1
(0018,9466)	Table X Position to Isocenter	FL	1
(0018,9467)	Table Y Position to Isocenter	FL	1
(0018,9468)	Table Z Position to Isocenter	FL	1
(0018,9469)	Table Horizontal Rotation Angle	FL	1
(0018,9470)	Table Head Tilt Angle	FL	1
(0018,9471)	Table Cradle Tilt Angle	FL	1
(0018,9472)	Frame Display Shutter Sequence	SQ	1
(0018,9473)	Acquired Image Area Dose Product	FL	1
(0018,9474)	C-arm Positioner Tabletop Relationship	CS	1

(0018,9476)	X-Ray Geometry Sequence	SQ	1
(0018,9477)	Irradiation Event Identification Sequence	SQ	1
(0018,A001)	Contributing Equipment Sequence	SQ	1
(0018,A002)	Contribution Date Time	DT	1
(0018,A003)	Contribution Description	ST	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(0020,000D)	Study Instance UID	UI	1	
(0020,000E)	Series Instance UID	UI	1	
(0020,0010)	Study ID	SH	1	
(0020,0011)	Series Number	IS	1	
(0020,0012)	Acquisition Number	IS	1	
(0020,0013)	Instance Number	IS	1	
(0020,0014)	<i>Isotope Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0015)	<i>Phase Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0016)	<i>Interval Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0017)	<i>Time Slot Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0018)	<i>Angle Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0019)	Item Number	IS	1	
(0020,0020)	Patient Orientation	CS	2	
(0020,0022)	<i>Overlay Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0024)	<i>Curve Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0026)	<i>Lookup Table Number</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,0030)	<i>Image Position</i>	<i>DS</i>	<i>3</i>	<i>RET</i>
(0020,0032)	Image Position (Patient)	DS	3	
(0020,0035)	<i>Image Orientation</i>	<i>DS</i>	<i>6</i>	<i>RET</i>
(0020,0037)	Image Orientation (Patient)	DS	6	
(0020,0050)	<i>Location</i>	<i>DS</i>	<i>1</i>	<i>RET</i>
(0020,0052)	Frame of Reference UID	UI	1	
(0020,0060)	Laterality	CS	1	
(0020,0062)	Image Laterality	CS	1	
(0020,0070)	<i>Image Geometry Type</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0020,0080)	<i>Masking Image</i>	<i>CS</i>	<i>1-n</i>	<i>RET</i>



(0020,0100)	Temporal Position Identifier	IS	1	
(0020,0105)	Number of Temporal Positions	IS	1	
(0020,0110)	Temporal Resolution	DS	1	
(0020,0200)	Synchronization Frame of Reference UID	UI	1	
(0020,1000)	<i>Series in Study</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,1001)	<i>Acquisitions in Series</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,1002)	Images in Acquisition	IS	1	
(0020,1003)	<i>Images in Series</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,1004)	<i>Acquisitions in Study</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,1005)	<i>Images in Study</i>	<i>IS</i>	<i>1</i>	<i>RET</i>
(0020,1020)	<i>Reference</i>	<i>CS</i>	<i>1-n</i>	<i>RET</i>
(0020,1040)	Position Reference Indicator	LO	1	
(0020,1041)	Slice Location	DS	1	
(0020,1070)	<i>Other Study Numbers</i>	<i>IS</i>	<i>1-n</i>	<i>RET</i>
(0020,1200)	Number of Patient Related Studies	IS	1	
(0020,1202)	Number of Patient Related Series	IS	1	
(0020,1204)	Number of Patient Related Instances	IS	1	
(0020,1206)	Number of Study Related Series	IS	1	
(0020,1208)	Number of Study Related Instances	IS	1	
(0020,1209)	Number of Series Related Instances	IS	1	
(0020,3100 to 31FF)	<i>Source Image IDs</i>	<i>CS</i>	<i>1-n</i>	<i>RET</i>
(0020,3401)	<i>Modifying Device ID</i>	<i>CS</i>	<i>1</i>	<i>RET</i>
(0020,3402)	<i>Modified Image ID</i>	<i>CS</i>	<i>1</i>	<i>RET</i>
(0020,3403)	<i>Modified Image Date</i>	<i>DA</i>	<i>1</i>	<i>RET</i>
(0020,3404)	<i>Modifying Device Manufacturer</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0020,3405)	<i>Modified Image Time</i>	<i>TM</i>	<i>1</i>	<i>RET</i>
(0020,3406)	<i>Modified Image Description</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0020,4000)	Image Comments	LT	1	
(0020,5000)	<i>Original Image Identification</i>	<i>AT</i>	<i>1-n</i>	<i>RET</i>
(0020,5002)	<i>Original Image Identification Nomenclature</i>	<i>CS</i>	<i>1-n</i>	<i>RET</i>
(0020,9056)	Stack ID	SH	1	
(0020,9057)	In-Stack Position Number	UL	1	
(0020,9071)	Frame Anatomy Sequence	SQ	1	

(0020,9072)	Frame Laterality	CS	1
(0020,9111)	Frame Content Sequence	SQ	1
(0020,9113)	Plane Position Sequence	SQ	1
(0020,9116)	Plane Orientation Sequence	SQ	1
(0020,9128)	Temporal Position Index	UL	1
(0020,9153)	Cardiac Trigger Delay Time	FD	1
(0020,9156)	Frame Acquisition Number	US	1
(0020,9157)	Dimension Index Values	UL	1-n
(0020,9158)	Frame Comments	LT	1
(0020,9161)	Concatenation UID	UI	1
(0020,9162)	In-concatenation Number	US	1
(0020,9163)	In-concatenation Total Number	US	1
(0020,9164)	Dimension Organization UID	UI	1
(0020,9165)	Dimension Index Pointer	AT	1
(0020,9167)	Functional Group Pointer	AT	1
(0020,9213)	Dimension Index Private Creator	LO	1
(0020,9221)	Dimension Organization Sequence	SQ	1
(0020,9222)	Dimension Index Sequence	SQ	1
(0020,9228)	Concatenation Frame Offset Number	UL	1
(0020,9238)	Functional Group Private Creator	LO	1
(0020,9251)	R – R Interval Time Measured	FD	1
(0020,9253)	Respiratory Trigger Sequence	SQ	1
(0020,9254)	Respiratory Interval Time	FD	1
(0020,9255)	Respiratory Trigger Delay Time	FD	1
(0020,9256)	Respiratory Trigger Delay Threshold	FD	1
(0020,9421)	Dimension Description Label	LO	1
(0020,9450)	Patient Orientation in Frame Sequence	SQ	1
(0020,9453)	Frame Label	LO	1

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0022,0001)	Light Path Filter Pass-Through Wavelength	US	1
(0022,0002)	Light Path Filter Pass Band	US	2
(0022,0003)	Image Path Filter Pass-Through Wavelength	US	1

(0022,0004)	Image Path Filter Pass Band	US	2
(0022,0005)	Patient Eye Movement Commanded	CS	1
(0022,0006)	Patient Eye Movement Command Code Sequence	SQ	1
(0022,0007)	Spherical Lens Power	FL	1
(0022,0008)	Cylinder Lens Power	FL	1
(0022,0009)	Cylinder Axis	FL	1
(0022,000A)	Emmetropic Magnification	FL	1
(0022,000B)	Intra Ocular Pressure	FL	1
(0022,000C)	Horizontal Field of View	FL	1
(0022,000D)	Pupil Dilated	CS	1
(0022,000E)	Degree of Dilation	FL	1
(0022,0010)	Stereo Baseline Angle	FL	1
(0022,0011)	Stereo Baseline Displacement	FL	1
(0022,0012)	Stereo Horizontal Pixel Offset	FL	1
(0022,0013)	Stereo Vertical Pixel Offset	FL	1
(0022,0014)	Stereo Rotation	FL	1
(0022,0015)	Acquisition Device Type Code Sequence	SQ	1
(0022,0016)	Illumination Type Code Sequence	SQ	1
(0022,0017)	Light Path Filter Type Stack Code Sequence	SQ	1
(0022,0018)	Image Path Filter Type Stack Code Sequence	SQ	1
(0022,0019)	Lenses Code Sequence	SQ	1
(0022,001A)	Channel Description Code Sequence	SQ	1
(0022,001B)	Refractive State Sequence	SQ	1
(0022,001C)	Mydriatic Agent Code Sequence	SQ	1
(0022,001D)	Relative Image Position Code Sequence	SQ	1
(0022,0020)	Stereo Pairs Sequence	SQ	1
(0022,0021)	Left Image Sequence	SQ	1
(0022,0022)	Right Image Sequence	SQ	1

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0028,0002)	Samples per Pixel	US	1
(0028,0003)	Samples per Pixel Used	US	1
(0028,0004)	Photometric Interpretation	CS	1

(0028,0005)	<i>Image Dimensions</i>	US	1	RET
(0028,0006)	Planar Configuration	US	1	
(0028,0008)	Number of Frames	IS	1	
(0028,0009)	Frame Increment Pointer	AT	1-n	
(0028,000A)	Frame Dimension Pointer	AT	1-n	
(0028,0010)	Rows	US	1	
(0028,0011)	Columns	US	1	
(0028,0012)	Planes	US	1	
(0028,0014)	Ultrasound Color Data Present	US	1	
(0028,0030)	Pixel Spacing	DS	2	
(0028,0031)	Zoom Factor	DS	2	
(0028,0032)	Zoom Center	DS	2	
(0028,0034)	Pixel Aspect Ratio	IS	2	
(0028,0040)	<i>Image Format</i>	CS	1	RET
(0028,0050)	<i>Manipulated Image</i>	LO	1-n	RET
(0028,0051)	Corrected Image	CS	1-n	
(0028,0060)	<i>Compression Code</i>	CS	1	RET
(0028,0100)	Bits Allocated	US	1	
(0028,0101)	Bits Stored	US	1	
(0028,0102)	High Bit	US	1	
(0028,0103)	Pixel Representation	US	1	
(0028,0104)	<i>Smallest Valid Pixel Value</i>	US or SS	1	RET
(0028,0105)	<i>Largest Valid Pixel Value</i>	US or SS	1	RET
(0028,0106)	Smallest Image Pixel Value	US or SS	1	
(0028,0107)	Largest Image Pixel Value	US or SS	1	
(0028,0108)	Smallest Pixel Value in Series	US or SS	1	
(0028,0109)	Largest Pixel Value in Series	US or SS	1	
(0028,0110)	Smallest Image Pixel Value in Plane	US or SS	1	
(0028,0111)	Largest Image Pixel Value in Plane	US or SS	1	
(0028,0120)	Pixel Padding Value	US or SS	1	
(0028,0200)	<i>Image Location</i>	US	1	RET
(0028,0300)	Quality Control Image	CS	1	
(0028,0301)	Burned In Annotation	CS	1	
(0028,0402)	Pixel Spacing Calibration Type	CS	1	

(0028,0404)	Pixel Spacing Calibration Description	LO	1	
(0028,1040)	Pixel Intensity Relationship	CS	1	
(0028,1041)	Pixel Intensity Relationship Sign	SS	1	
(0028,1050)	Window Center	DS	1-n	
(0028,1051)	Window Width	DS	1-n	
(0028,1052)	Rescale Intercept	DS	1	
(0028,1053)	Rescale Slope	DS	1	
(0028,1054)	Rescale Type	LO	1	
(0028,1055)	Window Center & Width Explanation	LO	1-n	
(0028,1056)	VOI LUT Function	CS	1	
(0028,1080)	<i>Gray Scale</i>	CS	1	<i>RET</i>
(0028,1090)	Recommended Viewing Mode	CS	1	
(0028,1100)	<i>Gray Lookup Table Descriptor</i>	<i>US or SS</i>	3	<i>RET</i>
(0028,1101)	Red Palette Color Lookup Table Descriptor	US or SS	3	
(0028,1102)	Green Palette Color Lookup Table Descriptor	US or SS	3	
(0028,1103)	Blue Palette Color Lookup Table Descriptor	US or SS	3	
(0028,1199)	Palette Color Lookup Table UID	UI	1	
(0028,1200)	<i>Gray Lookup Table Data</i>	<i>US or SS or OW</i>	<i>1-n 1</i>	<i>RET</i>
(0028,1201)	Red Palette Color Lookup Table Data	OW	1	
(0028,1202)	Green Palette Color Lookup Table Data	OW	1	
(0028,1203)	Blue Palette Color Lookup Table Data	OW	1	
(0028,1221)	Segmented Red Palette Color Lookup Table Data	OW	1	
(0028,1222)	Segmented Green Palette Color Lookup Table Data	OW	1	
(0028,1223)	Segmented Blue Palette Color Lookup Table Data	OW	1	
(0028,1300)	Implant Present	CS	1	
(0028,1350)	Partial View	CS	1	
(0028,1351)	Partial View Description	ST	1	
(0028,1352)	Partial View Code Sequence	SQ	1	
(0028,135A)	Spatial Locations Preserved	CS	1	
(0028,2000)	ICC Profile	OB	1	
(0028,2110)	Lossy Image Compression	CS	1	
(0028,2112)	Lossy Image Compression Ratio	DS	1-n	
(0028,2114)	Lossy Image Compression Method	CS	1-n	

(0028,3000)	Modality LUT Sequence	SQ	1	
(0028,3002)	LUT Descriptor	US or SS	3	
(0028,3003)	LUT Explanation	LO	1	
(0028,3004)	Modality LUT Type	LO	1	
(0028,3006)	LUT Data	US or SS or OW	1-n 1	
(0028,3010)	VOI LUT Sequence	SQ	1	
(0028,3110)	Softcopy VOI LUT Sequence	SQ	1	
(0028,4000)	<i>Image Presentation Comments</i>	<i>LT</i>	<i>1</i>	<i>RET</i>
(0028,5000)	Bi-Plane Acquisition Sequence	SQ	1	
(0028,6010)	Representative Frame Number	US	1	
(0028,6020)	Frame Numbers of Interest (FOI)	US	1-n	
(0028,6022)	Frame(s) of Interest Description	LO	1-n	
(0028,6023)	Frame of Interest Type	CS	1-n	
(0028,6030)	<i>Mask Pointer(s)</i>	<i>US</i>	<i>1-n</i>	<i>RET</i>
(0028,6040)	R Wave Pointer	US	1-n	
(0028,6100)	Mask Subtraction Sequence	SQ	1	
(0028,6101)	Mask Operation	CS	1	
(0028,6102)	Applicable Frame Range	US	2-2n	
(0028,6110)	Mask Frame Numbers	US	1-n	
(0028,6112)	Contrast Frame Averaging	US	1	
(0028,6114)	Mask Sub-pixel Shift	FL	2	
(0028,6120)	TID Offset	SS	1	
(0028,6190)	Mask Operation Explanation	ST	1	
(0028,7FE0)	Pixel Data Provider URL	UT	1	
(0028,9001)	Data Point Rows	UL	1	
(0028,9002)	Data Point Columns	UL	1	
(0028,9003)	Signal Domain Columns	CS	1	
(0028,9099)	<i>Largest Monochrome Pixel Value</i>	<i>US</i>	<i>1</i>	<i>RET</i>
(0028,9108)	Data Representation	CS	1	
(0028,9110)	Pixel Measures Sequence	SQ	1	
(0028,9132)	Frame VOI LUT Sequence	SQ	1	
(0028,9145)	Pixel Value Transformation Sequence	SQ	1	
(0028,9235)	Signal Domain Rows	CS	1	

(0028,9411)	Display Filter Percentage	FL	1
(0028,9415)	Frame Pixel Shift Sequence	SQ	1
(0028,9416)	Subtraction Item ID	US	1
(0028,9422)	Pixel Intensity Relationship LUT Sequence	SQ	1
(0028,9443)	Frame Pixel Data Properties Sequence	SQ	1
(0028,9444)	Geometrical Properties	CS	1
(0028,9445)	Geometric Maximum Distortion	FL	1
(0028,9446)	Image Processing Applied	CS	1-n
(0028,9454)	Mask Selection Mode	CS	1
(0028,9474)	LUT Function	CS	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(0032,000A)	<i>Study Status ID</i>	CS	1	RET
(0032,000C)	<i>Study Priority ID</i>	CS	1	RET
(0032,0012)	<i>Study ID Issuer</i>	LO	1	RET
(0032,0032)	<i>Study Verified Date</i>	DA	1	RET
(0032,0033)	<i>Study Verified Time</i>	TM	1	RET
(0032,0034)	<i>Study Read Date</i>	DA	1	RET
(0032,0035)	<i>Study Read Time</i>	TM	1	RET
(0032,1000)	<i>Scheduled Study Start Date</i>	DA	1	RET
(0032,1001)	<i>Scheduled Study Start Time</i>	TM	1	RET
(0032,1010)	<i>Scheduled Study Stop Date</i>	DA	1	RET
(0032,1011)	<i>Scheduled Study Stop Time</i>	TM	1	RET
(0032,1020)	<i>Scheduled Study Location</i>	LO	1	RET
(0032,1021)	<i>Scheduled Study Location AE Title</i>	AE	1-n	RET
(0032,1030)	<i>Reason for Study</i>	LO	1	RET
(0032,1031)	Requesting Physician Identification Sequence	SQ	1	
(0032,1032)	Requesting Physician	PN	1	
(0032,1033)	Requesting Service	LO	1	
(0032,1040)	<i>Study Arrival Date</i>	DA	1	RET
(0032,1041)	<i>Study Arrival Time</i>	TM	1	RET
(0032,1050)	<i>Study Completion Date</i>	DA	1	RET
(0032,1051)	<i>Study Completion Time</i>	TM	1	RET

(0032,1055)	<i>Study Component Status ID</i>	CS	1	<i>RET</i>
(0032,1060)	Requested Procedure Description	LO	1	
(0032,1064)	Requested Procedure Code Sequence	SQ	1	
(0032,1070)	Requested Contrast Agent	LO	1	
(0032,4000)	Study Comments	LT	1	

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(0038,0004)	Referenced Patient Alias Sequence	SQ	1	
(0038,0008)	Visit Status ID	CS	1	
(0038,0010)	Admission ID	LO	1	
(0038,0011)	Issuer of Admission ID	LO	1	
(0038,0016)	Route of Admissions	LO	1	
(0038,001A)	<i>Scheduled Admission Date</i>	<i>DA</i>	<i>1</i>	<i>RET</i>
(0038,001B)	<i>Scheduled Admission Time</i>	<i>TM</i>	<i>1</i>	<i>RET</i>
(0038,001C)	<i>Scheduled Discharge Date</i>	<i>DA</i>	<i>1</i>	<i>RET</i>
(0038,001D)	<i>Scheduled Discharge Time</i>	<i>TM</i>	<i>1</i>	<i>RET</i>
(0038,001E)	<i>Scheduled Patient Institution Residence</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0038,0020)	Admitting Date	DA	1	
(0038,0021)	Admitting Time	TM	1	
(0038,0030)	<i>Discharge Date</i>	<i>DA</i>	<i>1</i>	<i>RET</i>
(0038,0032)	<i>Discharge Time</i>	<i>TM</i>	<i>1</i>	<i>RET</i>
(0038,0040)	<i>Discharge Diagnosis Description</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0038,0044)	<i>Discharge Diagnosis Code Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
(0038,0050)	Special Needs	LO	1	
(0038,0100)	Pertinent Documents Sequence	SQ	1	
(0038,0300)	Current Patient Location	LO	1	
(0038,0400)	Patient's Institution Residence	LO	1	
(0038,0500)	Patient State	LO	1	
(0038,0502)	Patient Clinical Trial Participation Sequence	SQ	1	
(0038,4000)	Visit Comments	LT	1	

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(003A,0004)	Waveform Originality	CS	1	



(003A,0005)	Number of Waveform Channels	US	1
(003A,0010)	Number of Waveform Samples	UL	1
(003A,001A)	Sampling Frequency	DS	1
(003A,0020)	Multiplex Group Label	SH	1
(003A,0200)	Channel Definition Sequence	SQ	1
(003A,0202)	Waveform Channel Number	IS	1
(003A,0203)	Channel Label	SH	1
(003A,0205)	Channel Status	CS	1-n
(003A,0208)	Channel Source Sequence	SQ	1
(003A,0209)	Channel Source Modifiers Sequence	SQ	1
(003A,020A)	Source Waveform Sequence	SQ	1
(003A,020C)	Channel Derivation Description	LO	1
(003A,0210)	Channel Sensitivity	DS	1
(003A,0211)	Channel Sensitivity Units Sequence	SQ	1
(003A,0212)	Channel Sensitivity Correction Factor	DS	1
(003A,0213)	Channel Baseline	DS	1
(003A,0214)	Channel Time Skew	DS	1
(003A,0215)	Channel Sample Skew	DS	1
(003A,0218)	Channel Offset	DS	1
(003A,021A)	Waveform Bits Stored	US	1
(003A,0220)	Filter Low Frequency	DS	1
(003A,0221)	Filter High Frequency	DS	1
(003A,0222)	Notch Filter Frequency	DS	1
(003A,0223)	Notch Filter Bandwidth	DS	1
(003A,0300)	Multiplexed Audio Channels Description Code Sequence	SQ	1
(003A,0301)	Channel Identification Code	IS	1
(003A,0302)	Channel Mode	CS	1

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0040,0001)	Scheduled Station AE Title	AE	1-n
(0040,0002)	Scheduled Procedure Step Start Date	DA	1
(0040,0003)	Scheduled Procedure Step Start Time	TM	1
(0040,0004)	Scheduled Procedure Step End Date	DA	1

(0040,0005)	Scheduled Procedure Step End Time	TM	1
(0040,0006)	Scheduled Performing Physician's Name	PN	1
(0040,0007)	Scheduled Procedure Step Description	LO	1
(0040,0008)	Scheduled Protocol Code Sequence	SQ	1
(0040,0009)	Scheduled Procedure Step ID	SH	1
(0040,000A)	Stage Code Sequence	SQ	1
(0040,000B)	Scheduled Performing Physician Identification Sequence	SQ	1
(0040,0010)	Scheduled Station Name	SH	1-n
(0040,0011)	Scheduled Procedure Step Location	SH	1
(0040,0012)	Pre-Medication	LO	1
(0040,0020)	Scheduled Procedure Step Status	CS	1
(0040,0100)	Scheduled Procedure Step Sequence	SQ	1
(0040,0220)	Referenced Non-Image Composite SOP Instance Sequence	SQ	1
(0040,0241)	Performed Station AE Title	AE	1
(0040,0242)	Performed Station Name	SH	1
(0040,0243)	Performed Location	SH	1
(0040,0244)	Performed Procedure Step Start Date	DA	1
(0040,0245)	Performed Procedure Step Start Time	TM	1
(0040,0250)	Performed Procedure Step End Date	DA	1
(0040,0251)	Performed Procedure Step End Time	TM	1
(0040,0252)	Performed Procedure Step Status	CS	1
(0040,0253)	Performed Procedure Step ID	SH	1
(0040,0254)	Performed Procedure Step Description	LO	1
(0040,0255)	Performed Procedure Type Description	LO	1
(0040,0260)	Performed Protocol Code Sequence	SQ	1
(0040,0270)	Scheduled Step Attributes Sequence	SQ	1
(0040,0275)	Request Attributes Sequence	SQ	1
(0040,0280)	Comments on the Performed Procedure Step	ST	1
(0040,0281)	Performed Procedure Step Discontinuation Reason Code Sequence	SQ	1
(0040,0293)	Quantity Sequence	SQ	1
(0040,0294)	Quantity	DS	1
(0040,0295)	Measuring Units Sequence	SQ	1

(0040,0296)	Billing Item Sequence	SQ	1	
(0040,0300)	Total Time of Fluoroscopy	US	1	
(0040,0301)	Total Number of Exposures	US	1	
(0040,0302)	Entrance Dose	US	1	
(0040,0303)	Exposed Area	US	1-2	
(0040,0306)	Distance Source to Entrance	DS	1	
(0040,0307)	<i>Distance Source to Support</i>	<i>DS</i>	<i>1</i>	<i>RET</i>
(0040,030E)	Exposure Dose Sequence	SQ	1	
(0040,0310)	Comments on Radiation Dose	ST	1	
(0040,0312)	X-Ray Output	DS	1	
(0040,0314)	Half Value Layer	DS	1	
(0040,0316)	Organ Dose	DS	1	
(0040,0318)	Organ Exposed	CS	1	
(0040,0320)	Billing Procedure Step Sequence	SQ	1	
(0040,0321)	Film Consumption Sequence	SQ	1	
(0040,0324)	Billing Supplies and Devices Sequence	SQ	1	
(0040,0330)	<i>Referenced Procedure Step Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
(0040,0340)	Performed Series Sequence	SQ	1	
(0040,0400)	Comments on the Scheduled Procedure Step	LT	1	
(0040,0440)	Protocol Context Sequence	SQ	1	
(0040,0441)	Content Item Modifier Sequence	SQ	1	
(0040,050A)	Specimen Accession Number	LO	1	
(0040,0550)	Specimen Sequence	SQ	1	
(0040,0551)	Specimen Identifier	LO	1	
(0040,0555)	Acquisition Context Sequence	SQ	1	
(0040,0556)	Acquisition Context Description	ST	1	
(0040,059A)	Specimen Type Code Sequence	SQ	1	
(0040,06FA)	Slide Identifier	LO	1	
(0040,071A)	Image Center Point Coordinates Sequence	SQ	1	
(0040,072A)	X offset in Slide Coordinate System	DS	1	
(0040,073A)	Y offset in Slide Coordinate System	DS	1	
(0040,074A)	Z offset in Slide Coordinate System	DS	1	
(0040,08D8)	Pixel Spacing Sequence	SQ	1	
(0040,08DA)	Coordinate System Axis Code Sequence	SQ	1	

(0040,08EA)	Measurement Units Code Sequence	SQ	1	
(0040,1001)	Requested Procedure ID	SH	1	
(0040,1002)	Reason for the Requested Procedure	LO	1	
(0040,1003)	Requested Procedure Priority	SH	1	
(0040,1004)	Patient Transport Arrangements	LO	1	
(0040,1005)	Requested Procedure Location	LO	1	
(0040,1006)	<i>Placer Order Number / Procedure</i>	<i>SH</i>	<i>1</i>	<i>RET</i>
(0040,1007)	<i>Filler Order Number / Procedure</i>	<i>SH</i>	<i>1</i>	<i>RET</i>
(0040,1008)	Confidentiality Code	LO	1	
(0040,1009)	Reporting Priority	SH	1	
(0040,100A)	Reason for Requested Procedure Code Sequence	SQ	1	
(0040,1010)	Names of Intended Recipients of Results	PN	1-n	
(0040,1011)	Intended Recipients of Results Identification Sequence	SQ	1	
(0040,1101)	Person Identification Code Sequence	SQ	1	
(0040,1102)	Person's Address	ST	1	
(0040,1103)	Person's Telephone Numbers	LO	1-n	
(0040,1400)	Requested Procedure Comments	LT	1	
(0040,2001)	<i>Reason for the Imaging Service Request</i>	<i>LO</i>	<i>1</i>	<i>RET</i>
(0040,2004)	Issue Date of Imaging Service Request	DA	1	
(0040,2005)	Issue Time of Imaging Service Request	TM	1	
(0040,2006)	<i>Placer Order Number / Imaging Service Request</i>	<i>SH</i>	<i>1</i>	<i>RET</i>
(0040,2007)	<i>Filler Order Number / Imaging Service Request</i>	<i>SH</i>	<i>1</i>	<i>RET</i>
(0040,2008)	Order Entered By	PN	1	
(0040,2009)	Order Enterer's Location	SH	1	
(0040,2010)	Order Callback Phone Number	SH	1	
(0040,2016)	Placer Order Number / Imaging Service Request	LO	1	
(0040,2017)	Filler Order Number / Imaging Service Request	LO	1	
(0040,2400)	Imaging Service Request Comments	LT	1	
(0040,3001)	Confidentiality Constraint on Patient Data Description	LO	1	
(0040,4001)	General Purpose Scheduled Procedure Step Status	CS	1	
(0040,4002)	General Purpose Performed Procedure Step Status	CS	1	
(0040,4003)	General Purpose Scheduled Procedure Step Priority	CS	1	
(0040,4004)	Scheduled Processing Applications Code Sequence	SQ	1	
(0040,4005)	Scheduled Procedure Step Start Date and Time	DT	1	

(0040,4006)	Multiple Copies Flag	CS	1
(0040,4007)	Performed Processing Applications Code Sequence	SQ	1
(0040,4009)	Human Performer Code Sequence	SQ	1
(0040,4010)	Scheduled Procedure Step Modification Date and Time	DT	1
(0040,4011)	Expected Completion Date and Time	DT	1
(0040,4015)	Resulting General Purpose Performed Procedure Steps Sequence	SQ	1
(0040,4016)	Referenced General Purpose Scheduled Procedure Step Sequence	SQ	1
(0040,4018)	Scheduled Workitem Code Sequence	SQ	1
(0040,4019)	Performed Workitem Code Sequence	SQ	1
(0040,4020)	Input Availability Flag	CS	1
(0040,4021)	Input Information Sequence	SQ	1
(0040,4022)	Relevant Information Sequence	SQ	1
(0040,4023)	Referenced General Purpose Scheduled Procedure Step Transaction UID	UI	1
(0040,4025)	Scheduled Station Name Code Sequence	SQ	1
(0040,4026)	Scheduled Station Class Code Sequence	SQ	1
(0040,4027)	Scheduled Station Geographic Location Code Sequence	SQ	1
(0040,4028)	Performed Station Name Code Sequence	SQ	1
(0040,4029)	Performed Station Class Code Sequence	SQ	1
(0040,4030)	Performed Station Geographic Location Code Sequence	SQ	1
(0040,4031)	Requested Subsequent Workitem Code Sequence	SQ	1
(0040,4032)	Non-DICOM Output Code Sequence	SQ	1
(0040,4033)	Output Information Sequence	SQ	1
(0040,4034)	Scheduled Human Performers Sequence	SQ	1
(0040,4035)	Actual Human Performers Sequence	SQ	1
(0040,4036)	Human Performer's Organization	LO	1
(0040,4037)	Human Performer's Name	PN	1
(0040,8302)	Entrance Dose in mGy	DS	1
(0040,9094)	Referenced Image Real World Value Mapping Sequence	SQ	1
(0040,9096)	Real World Value Mapping Sequence	SQ	1
(0040,9098)	Pixel Value Mapping Code Sequence	SQ	1
(0040,9210)	LUT Label	SH	1
(0040,9211)	Real World Value Last Value Mapped	US or SS	1

(0040,9212)	Real World Value LUT Data	FD	1-n	
(0040,9216)	Real World Value First Value Mapped	US or SS	1	
(0040,9224)	Real World Value Intercept	FD	1	
(0040,9225)	Real World Value Slope	FD	1	
(0040,A010)	Relationship Type	CS	1	
(0040,A027)	Verifying Organization	LO	1	
(0040,A030)	Verification Date Time	DT	1	
(0040,A032)	Observation Date Time	DT	1	
(0040,A040)	Value Type	CS	1	
(0040,A043)	Concept Name Code Sequence	SQ	1	
(0040,A050)	Continuity Of Content	CS	1	
(0040,A073)	Verifying Observer Sequence	SQ	1	
(0040,A075)	Verifying Observer Name	PN	1	
(0040,A078)	Author Observer Sequence	SQ	1	
(0040,A07A)	Participant Sequence	SQ	1	
(0040,A07C)	Custodial Organization Sequence	SQ	1	
(0040,A080)	Participation Type	CS	1	
(0040,A082)	Participation Datetime	DT	1	
(0040,A084)	Observer Type	CS	1	
(0040,A088)	Verifying Observer Identification Code Sequence	SQ	1	
(0040,A090)	<i>Equivalent CDA Document Sequence</i>	SQ	1	<i>RET</i>
(0040,A0B0)	Referenced Waveform Channels	US	2-2n	
(0040,A120)	DateTime	DT	1	
(0040,A121)	Date	DA	1	
(0040,A122)	Time	TM	1	
(0040,A123)	Person Name	PN	1	
(0040,A124)	UID	UI	1	
(0040,A130)	Temporal Range Type	CS	1	
(0040,A132)	Referenced Sample Positions	UL	1-n	
(0040,A136)	Referenced Frame Numbers	US	1-n	
(0040,A138)	Referenced Time Offsets	DS	1-n	
(0040,A13A)	Referenced Datetime	DT	1-n	
(0040,A160)	Text Value	UT	1	
(0040,A168)	Concept Code Sequence	SQ	1	

(0040,A170)	Purpose of Reference Code Sequence	SQ	1	
(0040,A180)	Annotation Group Number	US	1	
(0040,A195)	Modifier Code Sequence	SQ	1	
(0040,A300)	Measured Value Sequence	SQ	1	
(0040,A301)	Numeric Value Qualifier Code Sequence	SQ	1	
(0040,A30A)	Numeric Value	DS	1-n	
(0040,A360)	Predecessor Documents Sequence	SQ	1	
(0040,A370)	Referenced Request Sequence	SQ	1	
(0040,A372)	Performed Procedure Code Sequence	SQ	1	
(0040,A375)	Current Requested Procedure Evidence Sequence	SQ	1	
(0040,A385)	Pertinent Other Evidence Sequence	SQ	1	
(0040,A390)	HL7 Structured Document Reference Sequence	SQ	1	
(0040,A491)	Completion Flag	CS	1	
(0040,A492)	Completion Flag Description	LO	1	
(0040,A493)	Verification Flag	CS	1	
(0040,A504)	Content Template Sequence	SQ	1	
(0040,A525)	Identical Documents Sequence	SQ	1	
(0040,A730)	Content Sequence	SQ	1	
(0040,B020)	Annotation Sequence	SQ	1	
(0040,DB00)	Template Identifier	CS	1	
<i>(0040,DB06)</i>	<i>Template Version</i>	<i>DT</i>	<i>1</i>	<i>RET</i>
<i>(0040,DB07)</i>	<i>Template Local Version</i>	<i>DT</i>	<i>1</i>	<i>RET</i>
<i>(0040,DB0B)</i>	<i>Template Extension Flag</i>	<i>CS</i>	<i>1</i>	<i>RET</i>
<i>(0040,DB0C)</i>	<i>Template Extension Organization UID</i>	<i>UI</i>	<i>1</i>	<i>RET</i>
<i>(0040,DB0D)</i>	<i>Template Extension Creator UID</i>	<i>UI</i>	<i>1</i>	<i>RET</i>
(0040,DB73)	Referenced Content Item Identifier	UL	1-n	
(0040,E001)	HL7 Instance Identifier	ST	1	
(0040,E004)	HL7 Document Effective Time	DT	1	
(0040,E006)	HL7 Document Type Code Sequence	SQ	1	
(0040,E010)	Retrieve URI	UT	1	

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0042,0010)	Document Title	ST	1

---

(0042,0011)	Encapsulated Document	OB	1
(0042,0012)	MIME Type of Encapsulated Document	LO	1
(0042,0013)	Source Instance Sequence	SQ	1

Tag	Name	VR	VM
(0050,0004)	Calibration Image	CS	1
(0050,0010)	Device Sequence	SQ	1
(0050,0014)	Device Length	DS	1
(0050,0016)	Device Diameter	DS	1
(0050,0017)	Device Diameter Units	CS	1
(0050,0018)	Device Volume	DS	1
(0050,0019)	Intermarker Distance	DS	1
(0050,0020)	Device Description	LO	1

Tag	Name	VR	VM
(0054,0010)	Energy Window Vector	US	1-n
(0054,0011)	Number of Energy Windows	US	1
(0054,0012)	Energy Window Information Sequence	SQ	1
(0054,0013)	Energy Window Range Sequence	SQ	1
(0054,0014)	Energy Window Lower Limit	DS	1
(0054,0015)	Energy Window Upper Limit	DS	1
(0054,0016)	Radiopharmaceutical Information Sequence	SQ	1
(0054,0017)	Residual Syringe Counts	IS	1
(0054,0018)	Energy Window Name	SH	1
(0054,0020)	Detector Vector	US	1-n
(0054,0021)	Number of Detectors	US	1
(0054,0022)	Detector Information Sequence	SQ	1
(0054,0030)	Phase Vector	US	1-n
(0054,0031)	Number of Phases	US	1
(0054,0032)	Phase Information Sequence	SQ	1
(0054,0033)	Number of Frames in Phase	US	1



(0054,0036)	Phase Delay	IS	1
(0054,0038)	Pause Between Frames	IS	1
(0054,0039)	Phase Description	CS	1
(0054,0050)	Rotation Vector	US	1-n
(0054,0051)	Number of Rotations	US	1
(0054,0052)	Rotation Information Sequence	SQ	1
(0054,0053)	Number of Frames in Rotation	US	1
(0054,0060)	R-R Interval Vector	US	1-n
(0054,0061)	Number of R-R Intervals	US	1
(0054,0062)	Gated Information Sequence	SQ	1
(0054,0063)	Data Information Sequence	SQ	1
(0054,0070)	Time Slot Vector	US	1-n
(0054,0071)	Number of Time Slots	US	1
(0054,0072)	Time Slot Information Sequence	SQ	1
(0054,0073)	Time Slot Time	DS	1
(0054,0080)	Slice Vector	US	1-n
(0054,0081)	Number of Slices	US	1
(0054,0090)	Angular View Vector	US	1-n
(0054,0100)	Time Slice Vector	US	1-n
(0054,0101)	Number of Time Slices	US	1
(0054,0200)	Start Angle	DS	1
(0054,0202)	Type of Detector Motion	CS	1
(0054,0210)	Trigger Vector	IS	1-n
(0054,0211)	Number of Triggers in Phase	US	1
(0054,0220)	View Code Sequence	SQ	1
(0054,0222)	View Modifier Code Sequence	SQ	1
(0054,0300)	Radionuclide Code Sequence	SQ	1
(0054,0302)	Administration Route Code Sequence	SQ	1
(0054,0304)	Radiopharmaceutical Code Sequence	SQ	1
(0054,0306)	Calibration Data Sequence	SQ	1
(0054,0308)	Energy Window Number	US	1
(0054,0400)	Image ID	SH	1
(0054,0410)	Patient Orientation Code Sequence	SQ	1
(0054,0412)	Patient Orientation Modifier Code Sequence	SQ	1

(0054,0414)	Patient Gantry Relationship Code Sequence	SQ	1
(0054,0500)	Slice Progression Direction	CS	1
(0054,1000)	Series Type	CS	2
(0054,1001)	Units	CS	1
(0054,1002)	Counts Source	CS	1
(0054,1004)	Reprojection Method	CS	1
(0054,1100)	Randoms Correction Method	CS	1
(0054,1101)	Attenuation Correction Method	LO	1
(0054,1102)	Decay Correction	CS	1
(0054,1103)	Reconstruction Method	LO	1
(0054,1104)	Detector Lines of Response Used	LO	1
(0054,1105)	Scatter Correction Method	LO	1
(0054,1200)	Axial Acceptance	DS	1
(0054,1201)	Axial Mash	IS	2
(0054,1202)	Transverse Mash	IS	1
(0054,1203)	Detector Element Size	DS	2
(0054,1210)	Coincidence Window Width	DS	1
(0054,1220)	Secondary Counts Type	CS	1-n
(0054,1300)	Frame Reference Time	DS	1
(0054,1310)	Primary (Prompts) Counts Accumulated	IS	1
(0054,1311)	Secondary Counts Accumulated	IS	1-n
(0054,1320)	Slice Sensitivity Factor	DS	1
(0054,1321)	Decay Factor	DS	1
(0054,1322)	Dose Calibration Factor	DS	1
(0054,1323)	Scatter Fraction Factor	DS	1
(0054,1324)	Dead Time Factor	DS	1
(0054,1330)	Image Index	US	1
(0054,1400)	Counts Included	CS	1-n
(0054,1401)	Dead Time Correction Flag	CS	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0060,3000)	Histogram Sequence	SQ	1
(0060,3002)	Histogram Number of Bins	US	1

(0060,3004)	Histogram First Bin Value	US or SS	1
(0060,3006)	Histogram Last Bin Value	US or SS	1
(0060,3008)	Histogram Bin Width	US	1
(0060,3010)	Histogram Explanation	LO	1
(0060,3020)	Histogram Data	UL	1-n

Tag	Name	VR	VM
(0062,0001)	Segmentation Type	CS	1
(0062,0002)	Segment Sequence	SQ	1
(0062,0003)	Segmented Property Category Code Sequence	SQ	1
(0062,0004)	Segment Number	US	1
(0062,0005)	Segment Label	LO	1
(0062,0006)	Segment Description	ST	1
(0062,0008)	Segment Algorithm Type	CS	1
(0062,0009)	Segment Algorithm Name	LO	1
(0062,000A)	Segment Identification Sequence	SQ	1
(0062,000B)	Referenced Segment Number	US	1-n
(0062,000C)	Recommended Display Grayscale Value	US	1
(0062,000D)	Recommended Display CIELab Value	US	3
(0062,000E)	Maximum Fractional Value	US	1
(0062,000F)	Segmented Property Type Code Sequence	SQ	1
(0062,0010)	Segmentation Fractional Type	CS	1

Tag	Name	VR	VM
(0064,0002)	Deformable Registration Sequence	SQ	1
(0064,0003)	Source Frame of Reference UID	UI	1
(0064,0005)	Deformable Registration Grid Sequence	SQ	1
(0064,0007)	Grid Dimensions	UL	3
(0064,0008)	Grid Resolution	FD	3
(0064,0009)	Vector Grid Data	OF	1
(0064,000F)	Pre Deformation Matrix Registration Sequence	SQ	1
(0064,0010)	Post Deformation Matrix Registration Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0070,0001)	Graphic Annotation Sequence	SQ	1
(0070,0002)	Graphic Layer	CS	1
(0070,0003)	Bounding Box Annotation Units	CS	1
(0070,0004)	Anchor Point Annotation Units	CS	1
(0070,0005)	Graphic Annotation Units	CS	1
(0070,0006)	Unformatted Text Value	ST	1
(0070,0008)	Text Object Sequence	SQ	1
(0070,0009)	Graphic Object Sequence	SQ	1
(0070,0010)	Bounding Box Top Left Hand Corner	FL	2
(0070,0011)	Bounding Box Bottom Right Hand Corner	FL	2
(0070,0012)	Bounding Box Text Horizontal Justification	CS	1
(0070,0014)	Anchor Point	FL	2
(0070,0015)	Anchor Point Visibility	CS	1
(0070,0020)	Graphic Dimensions	US	1
(0070,0021)	Number of Graphic Points	US	1
(0070,0022)	Graphic Data	FL	2-n
(0070,0023)	Graphic Type	CS	1
(0070,0024)	Graphic Filled	CS	1
(0070,0041)	Image Horizontal Flip	CS	1
(0070,0042)	Image Rotation	US	1
(0070,0052)	Displayed Area Top Left Hand Corner	SL	2
(0070,0053)	Displayed Area Bottom Right Hand Corner	SL	2
(0070,005A)	Displayed Area Selection Sequence	SQ	1
(0070,0060)	Graphic Layer Sequence	SQ	1
(0070,0062)	Graphic Layer Order	IS	1
(0070,0066)	Graphic Layer Recommended Display Grayscale Value	US	1
(0070,0067)	<i>Graphic Layer Recommended Display RGB Value</i>	US	3 <i>RET</i>
(0070,0068)	Graphic Layer Description	LO	1
(0070,0080)	Content Label	CS	1
(0070,0081)	Content Description	LO	1
(0070,0082)	Presentation Creation Date	DA	1

(0070,0083)	Presentation Creation Time	TM	1
(0070,0084)	Content Creator's Name	PN	1
(0070,0086)	Content Creator's Identification Code Sequence	SQ	1
(0070,0100)	Presentation Size Mode	CS	1
(0070,0101)	Presentation Pixel Spacing	DS	2
(0070,0102)	Presentation Pixel Aspect Ratio	IS	2
(0070,0103)	Presentation Pixel Magnification Ratio	FL	1
(0070,0306)	Shape Type	CS	1
(0070,0308)	Registration Sequence	SQ	1
(0070,0309)	Matrix Registration Sequence	SQ	1
(0070,030A)	Matrix Sequence	SQ	1
(0070,030C)	Frame of Reference Transformation Matrix Type	CS	1
(0070,030D)	Registration Type Code Sequence	SQ	1
(0070,030F)	Fiducial Description	ST	1
(0070,0310)	Fiducial Identifier	SH	1
(0070,0311)	Fiducial Identifier Code Sequence	SQ	1
(0070,0312)	Contour Uncertainty Radius	FD	1
(0070,0314)	Used Fiducials Sequence	SQ	1
(0070,0318)	Graphic Coordinates Data Sequence	SQ	1
(0070,031A)	Fiducial UID	UI	1
(0070,031C)	Fiducial Set Sequence	SQ	1
(0070,031E)	Fiducial Sequence	SQ	1
(0070,0401)	Graphic Layer Recommended Display CIELab Value	US	3
(0070,0402)	Blending Sequence	SQ	1
(0070,0403)	Relative Opacity	FL	1
(0070,0404)	Referenced Spatial Registration Sequence	SQ	1
(0070,0405)	Blending Position	CS	1

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0072,0002)	Hanging Protocol Name	SH	1
(0072,0004)	Hanging Protocol Description	LO	1
(0072,0006)	Hanging Protocol Level	CS	1
(0072,0008)	Hanging Protocol Creator	LO	1

(0072,000A)	Hanging Protocol Creation Datetime	DT	1
(0072,000C)	Hanging Protocol Definition Sequence	SQ	1
(0072,000E)	Hanging Protocol User Identification Code Sequence	SQ	1
(0072,0010)	Hanging Protocol User Group Name	LO	1
(0072,0012)	Source Hanging Protocol Sequence	SQ	1
(0072,0014)	Number of Priors Referenced	US	1
(0072,0020)	Image Sets Sequence	SQ	1
(0072,0022)	Image Set Selector Sequence	SQ	1
(0072,0024)	Image Set Selector Usage Flag	CS	1
(0072,0026)	Selector Attribute	AT	1
(0072,0028)	Selector Value Number	US	1
(0072,0030)	Time Based Image Sets Sequence	SQ	1
(0072,0032)	Image Set Number	US	1
(0072,0034)	Image Set Selector Category	CS	1
(0072,0038)	Relative Time	US	2
(0072,003A)	Relative Time Units	CS	1
(0072,003C)	Abstract Prior Value	SS	2
(0072,003E)	Abstract Prior Code Sequence	SQ	1
(0072,0040)	Image Set Label	LO	1
(0072,0050)	Selector Attribute VR	CS	1
(0072,0052)	Selector Sequence Pointer	AT	1
(0072,0054)	Selector Sequence Pointer Private Creator	LO	1
(0072,0056)	Selector Attribute Private Creator	LO	1
(0072,0060)	Selector AT Value	AT	1-n
(0072,0062)	Selector CS Value	CS	1-n
(0072,0064)	Selector IS Value	IS	1-n
(0072,0066)	Selector LO Value	LO	1-n
(0072,0068)	Selector LT Value	LT	1-n
(0072,006A)	Selector PN Value	PN	1-n
(0072,006C)	Selector SH Value	SH	1-n
(0072,006E)	Selector ST Value	ST	1-n
(0072,0070)	Selector UT Value	UT	1-n
(0072,0072)	Selector DS Value	DS	1-n
(0072,0074)	Selector FD Value	FD	1-n

(0072,0076)	Selector FL Value	FL	1-n
(0072,0078)	Selector UL Value	UL	1-n
(0072,007A)	Selector US Value	US	1-n
(0072,007C)	Selector SL Value	SL	1-n
(0072,007E)	Selector SS Value	SS	1-n
(0072,0080)	Selector Code Sequence Value	SQ	1
(0072,0100)	Number of Screens	US	1
(0072,0102)	Nominal Screen Definition Sequence	SQ	1
(0072,0104)	Number of Vertical Pixels	US	1
(0072,0106)	Number of Horizontal Pixels	US	1
(0072,0108)	Display Environment Spatial Position	FD	4
(0072,010A)	Screen Minimum Grayscale Bit Depth	US	1
(0072,010C)	Screen Minimum Color Bit Depth	US	1
(0072,010E)	Application Maximum Repaint Time	US	1
(0072,0200)	Display Sets Sequence	SQ	1
(0072,0202)	Display Set Number	US	1
(0072,0203)	Display Set Label	LO	1
(0072,0204)	Display Set Presentation Group	US	1
(0072,0206)	Display Set Presentation Group Description	LO	1
(0072,0208)	Partial Data Display Handling	CS	1
(0072,0210)	Synchronized Scrolling Sequence	SQ	1
(0072,0212)	Display Set Scrolling Group	US	2-n
(0072,0214)	Navigation Indicator Sequence	SQ	1
(0072,0216)	Navigation Display Set	US	1
(0072,0218)	Reference Display Sets	US	1-n
(0072,0300)	Image Boxes Sequence	SQ	1
(0072,0302)	Image Box Number	US	1
(0072,0304)	Image Box Layout Type	CS	1
(0072,0306)	Image Box Tile Horizontal Dimension	US	1
(0072,0308)	Image Box Tile Vertical Dimension	US	1
(0072,0310)	Image Box Scroll Direction	CS	1
(0072,0312)	Image Box Small Scroll Type	CS	1
(0072,0314)	Image Box Small Scroll Amount	US	1
(0072,0316)	Image Box Large Scroll Type	CS	1

(0072,0318)	Image Box Large Scroll Amount	US	1
(0072,0320)	Image Box Overlap Priority	US	1
(0072,0330)	Cine Relative to Real-Time	FD	1
(0072,0400)	Filter Operations Sequence	SQ	1
(0072,0402)	Filter-by Category	CS	1
(0072,0404)	Filter-by Attribute Presence	CS	1
(0072,0406)	Filter-by Operator	CS	1
(0072,0500)	Blending Operation Type	CS	1
(0072,0510)	Reformatting Operation Type	CS	1
(0072,0512)	Reformatting Thickness	FD	1
(0072,0514)	Reformatting Interval	FD	1
(0072,0516)	Reformatting Operation Initial View Direction	CS	1
(0072,0520)	3D Rendering Type	CS	1-n
(0072,0600)	Sorting Operations Sequence	SQ	1
(0072,0602)	Sort-by Category	CS	1
(0072,0604)	Sorting Direction	CS	1
(0072,0700)	Display Set Patient Orientation	CS	2
(0072,0702)	VOI Type	CS	1
(0072,0704)	Pseudo-color Type	CS	1
(0072,0706)	Show Grayscale Inverted	CS	1
(0072,0710)	Show Image True Size Flag	CS	1
(0072,0712)	Show Graphic Annotation Flag	CS	1
(0072,0714)	Show Patient Demographics Flag	CS	1
(0072,0716)	Show Acquisition Techniques Flag	CS	1
(0072,0717)	Display Set Horizontal Justification	CS	1
(0072,0718)	Display Set Vertical Justification	CS	1

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0088,0130)	Storage Media File-set ID	SH	1
(0088,0140)	Storage Media File-set UID	UI	1
(0088,0200)	Icon Image Sequence	SQ	1
(0088,0904)	Topic Title	LO	1
(0088,0906)	Topic Subject	ST	1



(0088,0910)	Topic Author	LO	1
(0088,0912)	Topic Keywords	LO	1-32

Tag	Name	VR	VM
(0100,0410)	SOP Instance Status	CS	1
(0100,0420)	SOP Authorization Date and Time	DT	1
(0100,0424)	SOP Authorization Comment	LT	1
(0100,0426)	Authorization Equipment Certification Number	LO	1

Tag	Name	VR	VM
(0400,0005)	MAC ID Number	US	1
(0400,0010)	MAC Calculation Transfer Syntax UID	UI	1
(0400,0015)	MAC Algorithm	CS	1
(0400,0020)	Data Elements Signed	AT	1-n
(0400,0100)	Digital Signature UID	UI	1
(0400,0105)	Digital Signature DateTime	DT	1
(0400,0110)	Certificate Type	CS	1
(0400,0115)	Certificate of Signer	OB	1
(0400,0120)	Signature	OB	1
(0400,0305)	Certified Timestamp Type	CS	1
(0400,0310)	Certified Timestamp	OB	1
(0400,0401)	Digital Signature Purpose Code Sequence	SQ	1
(0400,0402)	Referenced Digital Signature Sequence	SQ	1
(0400,0403)	Referenced SOP Instance MAC Sequence	SQ	1
(0400,0404)	MAC	OB	1
(0400,0500)	Encrypted Attributes Sequence	SQ	1
(0400,0510)	Encrypted Content Transfer Syntax UID	UI	1
(0400,0520)	Encrypted Content	OB	1
(0400,0550)	Modified Attributes Sequence	SQ	1
(0400,0561)	Original Attributes Sequence	SQ	1
(0400,0562)	Attribute Modification Datetime	DT	1
(0400,0563)	Modifying System	LO	1
(0400,0564)	Source of Previous Values	LO	1

(0400,0565)	Reason for the Attribute Modification	CS	1
-------------	---------------------------------------	----	---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(2000,0010)	Number of Copies	IS	1
(2000,001E)	Printer Configuration Sequence	SQ	1
(2000,0020)	Print Priority	CS	1
(2000,0030)	Medium Type	CS	1
(2000,0040)	Film Destination	CS	1
(2000,0050)	Film Session Label	LO	1
(2000,0060)	Memory Allocation	IS	1
(2000,0061)	Maximum Memory Allocation	IS	1
(2000,0062)	<i>Color Image Printing Flag</i>	CS	1 <i>RET</i>
(2000,0063)	<i>Collation Flag</i>	CS	1 <i>RET</i>
(2000,0065)	<i>Annotation Flag</i>	CS	1 <i>RET</i>
(2000,0067)	<i>Image Overlay Flag</i>	CS	1 <i>RET</i>
(2000,0069)	<i>Presentation LUT Flag</i>	CS	1 <i>RET</i>
(2000,006A)	<i>Image Box Presentation LUT Flag</i>	CS	1 <i>RET</i>
(2000,00A0)	Memory Bit Depth	US	1
(2000,00A1)	Printing Bit Depth	US	1
(2000,00A2)	Media Installed Sequence	SQ	1
(2000,00A4)	Other Media Available Sequence	SQ	1
(2000,00A8)	Supported Image Display Formats Sequence	SQ	1
(2000,0500)	Referenced Film Box Sequence	SQ	1
(2000,0510)	Referenced Stored Print Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(2010,0010)	Image Display Format	ST	1
(2010,0030)	Annotation Display Format ID	CS	1
(2010,0040)	Film Orientation	CS	1
(2010,0050)	Film Size ID	CS	1
(2010,0052)	Printer Resolution ID	CS	1

(2010,0054)	Default Printer Resolution ID	CS	1
(2010,0060)	Magnification Type	CS	1
(2010,0080)	Smoothing Type	CS	1
(2010,00A6)	Default Magnification Type	CS	1
(2010,00A7)	Other Magnification Types Available	CS	1-n
(2010,00A8)	Default Smoothing Type	CS	1
(2010,00A9)	Other Smoothing Types Available	CS	1-n
(2010,0100)	Border Density	CS	1
(2010,0110)	Empty Image Density	CS	1
(2010,0120)	Min Density	US	1
(2010,0130)	Max Density	US	1
(2010,0140)	Trim	CS	1
(2010,0150)	Configuration Information	ST	1
(2010,0152)	Configuration Information Description	LT	1
(2010,0154)	Maximum Collated Films	IS	1
(2010,015E)	Illumination	US	1
(2010,0160)	Reflected Ambient Light	US	1
(2010,0376)	Printer Pixel Spacing	DS	2
(2010,0500)	Referenced Film Session Sequence	SQ	1
(2010,0510)	Referenced Image Box Sequence	SQ	1
(2010,0520)	Referenced Basic Annotation Box Sequence	SQ	1

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(2020,0010)	Image Position	US	1
(2020,0020)	Polarity	CS	1
(2020,0030)	Requested Image Size	DS	1
(2020,0040)	Requested Decimate/Crop Behavior	CS	1
(2020,0050)	Requested Resolution ID	CS	1
(2020,00A0)	Requested Image Size Flag	CS	1
(2020,00A2)	Decimate/Crop Result	CS	1
(2020,0110)	Basic Grayscale Image Sequence	SQ	1
(2020,0111)	Basic Color Image Sequence	SQ	1

(2020,0130)	Referenced Image Overlay Box Sequence	SQ	1	RET
(2020,0140)	Referenced VOI LUT Box Sequence	SQ	1	RET

Tag	Name	VR	VM
(2030,0010)	Annotation Position	US	1
(2030,0020)	Text String	LO	1

Tag	Name	VR	VM
(2040,0010)	Referenced Overlay Plane Sequence	SQ	1 RET
(2040,0011)	Referenced Overlay Plane Groups	US	1-99 RET
(2040,0020)	Overlay Pixel Data Sequence	SQ	1 RET
(2040,0060)	Overlay Magnification Type	CS	1 RET
(2040,0070)	Overlay Smoothing Type	CS	1 RET
(2040,0072)	Overlay or Image Magnification	CS	1 RET
(2040,0074)	Magnify to Number of Columns	US	1 RET
(2040,0080)	Overlay Foreground Density	CS	1 RET
(2040,0082)	Overlay Background Density	CS	1 RET
(2040,0090)	Overlay Mode	CS	1 RET
(2040,0100)	Threshold Density	CS	1 RET
(2040,0500)	Referenced Image Box Sequence	SQ	1 RET

Tag	Name	VR	VM
(2050,0010)	Presentation LUT Sequence	SQ	1
(2050,0020)	Presentation LUT Shape	CS	1
(2050,0500)	Referenced Presentation LUT Sequence	SQ	1

Tag	Name	VR	VM
(2100,0010)	Print Job ID	SH	1
(2100,0020)	Execution Status	CS	1
(2100,0030)	Execution Status Info	CS	1
(2100,0040)	Creation Date	DA	1

(2100,0050)	Creation Time	TM	1	
(2100,0070)	Originator	AE	1	
(2100,0140)	Destination AE	AE	1	
(2100,0160)	Owner ID	SH	1	
(2100,0170)	Number of Films	IS	1	
<i>(2100,0500)</i>	<i>Referenced Print Job Sequence (Pull Stored Print)</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(2110,0010)	Printer Status	CS	1	
(2110,0020)	Printer Status Info	CS	1	
(2110,0030)	Printer Name	LO	1	
<i>(2110,0099)</i>	<i>Print Queue ID</i>	<i>SH</i>	<i>1</i>	<i>RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
<i>(2120,0010)</i>	<i>Queue Status</i>	<i>CS</i>	<i>1</i>	<i>RET</i>
<i>(2120,0050)</i>	<i>Print Job Description Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2120,0070)</i>	<i>Referenced Print Job Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
<i>(2130,0010)</i>	<i>Print Management Capabilities Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0015)</i>	<i>Printer Characteristics Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0030)</i>	<i>Film Box Content Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0040)</i>	<i>Image Box Content Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0050)</i>	<i>Annotation Content Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0060)</i>	<i>Image Overlay Box Content Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,0080)</i>	<i>Presentation LUT Content Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,00A0)</i>	<i>Proposed Study Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>
<i>(2130,00C0)</i>	<i>Original Image Sequence</i>	<i>SQ</i>	<i>1</i>	<i>RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(2200,0001)	Label Using Information Extracted From Instances	CS	1	
(2200,0002)	Label Text	UT	1	

(2200,0003)	Label Style Selection	CS	1
(2200,0004)	Media Disposition	LT	1
(2200,0005)	Barcode Value	LT	1
(2200,0006)	Barcode Symbology	CS	1
(2200,0007)	Allow Media Splitting	CS	1
(2200,0008)	Include Non-DICOM Objects	CS	1
(2200,0009)	Include Display Application	CS	1
(2200,000A)	Preserve Composite Instances After Media Creation	CS	1
(2200,000B)	Total Number of Pieces of Media Created	US	1
(2200,000C)	Requested Media Application Profile	LO	1
(2200,000D)	Referenced Storage Media Sequence	SQ	1
(2200,000E)	Failure Attributes	AT	1-n
(2200,000F)	Allow Lossy Compression	CS	1
(2200,0020)	Request Priority	CS	1

---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(3002,0002)	RT Image Label	SH	1
(3002,0003)	RT Image Name	LO	1
(3002,0004)	RT Image Description	ST	1
(3002,000A)	Reported Values Origin	CS	1
(3002,000C)	RT Image Plane	CS	1
(3002,000D)	X-Ray Image Receptor Translation	DS	3
(3002,000E)	X-Ray Image Receptor Angle	DS	1
(3002,0010)	RT Image Orientation	DS	6
(3002,0011)	Image Plane Pixel Spacing	DS	2
(3002,0012)	RT Image Position	DS	2
(3002,0020)	Radiation Machine Name	SH	1
(3002,0022)	Radiation Machine SAD	DS	1
(3002,0024)	Radiation Machine SSD	DS	1
(3002,0026)	RT Image SID	DS	1
(3002,0028)	Source to Reference Object Distance	DS	1
(3002,0029)	Fraction Number	IS	1
(3002,0030)	Exposure Sequence	SQ	1

(3002,0032)	Meterset Exposure	DS	1
(3002,0034)	Diaphragm Position	DS	4
(3002,0040)	Fluence Map Sequence	SQ	1
(3002,0041)	Fluence Data Source	CS	1
(3002,0042)	Fluence Data Scale	DS	1

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(3004,0001)	DVH Type	CS	1
(3004,0002)	Dose Units	CS	1
(3004,0004)	Dose Type	CS	1
(3004,0006)	Dose Comment	LO	1
(3004,0008)	Normalization Point	DS	3
(3004,000A)	Dose Summation Type	CS	1
(3004,000C)	Grid Frame Offset Vector	DS	2-n
(3004,000E)	Dose Grid Scaling	DS	1
(3004,0010)	RT Dose ROI Sequence	SQ	1
(3004,0012)	Dose Value	DS	1
(3004,0014)	Tissue Heterogeneity Correction	CS	1-3
(3004,0040)	DVH Normalization Point	DS	3
(3004,0042)	DVH Normalization Dose Value	DS	1
(3004,0050)	DVH Sequence	SQ	1
(3004,0052)	DVH Dose Scaling	DS	1
(3004,0054)	DVH Volume Units	CS	1
(3004,0056)	DVH Number of Bins	IS	1
(3004,0058)	DVH Data	DS	2-2n
(3004,0060)	DVH Referenced ROI Sequence	SQ	1
(3004,0062)	DVH ROI Contribution Type	CS	1
(3004,0070)	DVH Minimum Dose	DS	1
(3004,0072)	DVH Maximum Dose	DS	1
(3004,0074)	DVH Mean Dose	DS	1

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(3006,0002)	Structure Set Label	SH	1

(3006,0004)	Structure Set Name	LO	1
(3006,0006)	Structure Set Description	ST	1
(3006,0008)	Structure Set Date	DA	1
(3006,0009)	Structure Set Time	TM	1
(3006,0010)	Referenced Frame of Reference Sequence	SQ	1
(3006,0012)	RT Referenced Study Sequence	SQ	1
(3006,0014)	RT Referenced Series Sequence	SQ	1
(3006,0016)	Contour Image Sequence	SQ	1
(3006,0020)	Structure Set ROI Sequence	SQ	1
(3006,0022)	ROI Number	IS	1
(3006,0024)	Referenced Frame of Reference UID	UI	1
(3006,0026)	ROI Name	LO	1
(3006,0028)	ROI Description	ST	1
(3006,002A)	ROI Display Color	IS	3
(3006,002C)	ROI Volume	DS	1
(3006,0030)	RT Related ROI Sequence	SQ	1
(3006,0033)	RT ROI Relationship	CS	1
(3006,0036)	ROI Generation Algorithm	CS	1
(3006,0038)	ROI Generation Description	LO	1
(3006,0039)	ROI Contour Sequence	SQ	1
(3006,0040)	Contour Sequence	SQ	1
(3006,0042)	Contour Geometric Type	CS	1
(3006,0044)	Contour Slab Thickness	DS	1
(3006,0045)	Contour Offset Vector	DS	3
(3006,0046)	Number of Contour Points	IS	1
(3006,0048)	Contour Number	IS	1
(3006,0049)	Attached Contours	IS	1-n
(3006,0050)	Contour Data	DS	3-3n
(3006,0080)	RT ROI Observations Sequence	SQ	1
(3006,0082)	Observation Number	IS	1
(3006,0084)	Referenced ROI Number	IS	1
(3006,0085)	ROI Observation Label	SH	1
(3006,0086)	RT ROI Identification Code Sequence	SQ	1
(3006,0088)	ROI Observation Description	ST	1



(3006,00A0)	Related RT ROI Observations Sequence	SQ	1
(3006,00A4)	RT ROI Interpreted Type	CS	1
(3006,00A6)	ROI Interpreter	PN	1
(3006,00B0)	ROI Physical Properties Sequence	SQ	1
(3006,00B2)	ROI Physical Property	CS	1
(3006,00B4)	ROI Physical Property Value	DS	1
(3006,00C0)	Frame of Reference Relationship Sequence	SQ	1
(3006,00C2)	Related Frame of Reference UID	UI	1
(3006,00C4)	Frame of Reference Transformation Type	CS	1
(3006,00C6)	Frame of Reference Transformation Matrix	DS	16
(3006,00C8)	Frame of Reference Transformation Comment	LO	1

---



---

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(3008,0010)	Measured Dose Reference Sequence	SQ	1
(3008,0012)	Measured Dose Description	ST	1
(3008,0014)	Measured Dose Type	CS	1
(3008,0016)	Measured Dose Value	DS	1
(3008,0020)	Treatment Session Beam Sequence	SQ	1
(3008,0021)	Treatment Session Ion Beam Sequence	SQ	1
(3008,0022)	Current Fraction Number	IS	1
(3008,0024)	Treatment Control Point Date	DA	1
(3008,0025)	Treatment Control Point Time	TM	1
(3008,002A)	Treatment Termination Status	CS	1
(3008,002B)	Treatment Termination Code	SH	1
(3008,002C)	Treatment Verification Status	CS	1
(3008,0030)	Referenced Treatment Record Sequence	SQ	1
(3008,0032)	Specified Primary Meterset	DS	1
(3008,0033)	Specified Secondary Meterset	DS	1
(3008,0036)	Delivered Primary Meterset	DS	1
(3008,0037)	Delivered Secondary Meterset	DS	1
(3008,003A)	Specified Treatment Time	DS	1
(3008,003B)	Delivered Treatment Time	DS	1

(3008,0040)	Control Point Delivery Sequence	SQ	1
(3008,0041)	Ion Control Point Delivery Sequence	SQ	1
(3008,0042)	Specified Meterset	DS	1
(3008,0044)	Delivered Meterset	DS	1
(3008,0045)	Meterset Rate Set	FL	1
(3008,0046)	Meterset Rate Delivered	FL	1
(3008,0047)	Scan Spot Metersets Delivered	FL	1-n
(3008,0048)	Dose Rate Delivered	DS	1
(3008,0050)	Treatment Summary Calculated Dose Reference Sequence	SQ	1
(3008,0052)	Cumulative Dose to Dose Reference	DS	1
(3008,0054)	First Treatment Date	DA	1
(3008,0056)	Most Recent Treatment Date	DA	1
(3008,005A)	Number of Fractions Delivered	IS	1
(3008,0060)	Override Sequence	SQ	1
(3008,0061)	Parameter Sequence Pointer	AT	1
(3008,0062)	Override Parameter Pointer	AT	1
(3008,0063)	Parameter Item Index	IS	1
(3008,0064)	Measured Dose Reference Number	IS	1
(3008,0065)	Parameter Pointer	AT	1
(3008,0066)	Override Reason	ST	1
(3008,0068)	Corrected Parameter Sequence	SQ	1
(3008,006A)	Correction Value	FL	1
(3008,0070)	Calculated Dose Reference Sequence	SQ	1
(3008,0072)	Calculated Dose Reference Number	IS	1
(3008,0074)	Calculated Dose Reference Description	ST	1
(3008,0076)	Calculated Dose Reference Dose Value	DS	1
(3008,0078)	Start Meterset	DS	1
(3008,007A)	End Meterset	DS	1
(3008,0080)	Referenced Measured Dose Reference Sequence	SQ	1
(3008,0082)	Referenced Measured Dose Reference Number	IS	1
(3008,0090)	Referenced Calculated Dose Reference Sequence	SQ	1
(3008,0092)	Referenced Calculated Dose Reference Number	IS	1
(3008,00A0)	Beam Limiting Device Leaf Pairs Sequence	SQ	1

(3008,00B0)	Recorded Wedge Sequence	SQ	1
(3008,00C0)	Recorded Compensator Sequence	SQ	1
(3008,00D0)	Recorded Block Sequence	SQ	1
(3008,00E0)	Treatment Summary Measured Dose Reference Sequence	SQ	1
(3008,00F0)	Recorded Snout Sequence	SQ	1
(3008,00F2)	Recorded Range Shifter Sequence	SQ	1
(3008,00F4)	Recorded Lateral Spreading Device Sequence	SQ	1
(3008,00F6)	Recorded Range Modulator Sequence	SQ	1
(3008,0100)	Recorded Source Sequence	SQ	1
(3008,0105)	Source Serial Number	LO	1
(3008,0110)	Treatment Session Application Setup Sequence	SQ	1
(3008,0116)	Application Setup Check	CS	1
(3008,0120)	Recorded Brachy Accessory Device Sequence	SQ	1
(3008,0122)	Referenced Brachy Accessory Device Number	IS	1
(3008,0130)	Recorded Channel Sequence	SQ	1
(3008,0132)	Specified Channel Total Time	DS	1
(3008,0134)	Delivered Channel Total Time	DS	1
(3008,0136)	Specified Number of Pulses	IS	1
(3008,0138)	Delivered Number of Pulses	IS	1
(3008,013A)	Specified Pulse Repetition Interval	DS	1
(3008,013C)	Delivered Pulse Repetition Interval	DS	1
(3008,0140)	Recorded Source Applicator Sequence	SQ	1
(3008,0142)	Referenced Source Applicator Number	IS	1
(3008,0150)	Recorded Channel Shield Sequence	SQ	1
(3008,0152)	Referenced Channel Shield Number	IS	1
(3008,0160)	Brachy Control Point Delivered Sequence	SQ	1
(3008,0162)	Safe Position Exit Date	DA	1
(3008,0164)	Safe Position Exit Time	TM	1
(3008,0166)	Safe Position Return Date	DA	1
(3008,0168)	Safe Position Return Time	TM	1
(3008,0200)	Current Treatment Status	CS	1
(3008,0202)	Treatment Status Comment	ST	1
(3008,0220)	Fraction Group Summary Sequence	SQ	1

(3008,0223)	Referenced Fraction Number	IS	1
(3008,0224)	Fraction Group Type	CS	1
(3008,0230)	Beam Stopper Position	CS	1
(3008,0240)	Fraction Status Summary Sequence	SQ	1
(3008,0250)	Treatment Date	DA	1
(3008,0251)	Treatment Time	TM	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(300A,0002)	RT Plan Label	SH	1
(300A,0003)	RT Plan Name	LO	1
(300A,0004)	RT Plan Description	ST	1
(300A,0006)	RT Plan Date	DA	1
(300A,0007)	RT Plan Time	TM	1
(300A,0009)	Treatment Protocols	LO	1-n
(300A,000A)	Plan Intent	CS	1
(300A,000B)	Treatment Sites	LO	1-n
(300A,000C)	RT Plan Geometry	CS	1
(300A,000E)	Prescription Description	ST	1
(300A,0010)	Dose Reference Sequence	SQ	1
(300A,0012)	Dose Reference Number	IS	1
(300A,0013)	Dose Reference UID	UI	1
(300A,0014)	Dose Reference Structure Type	CS	1
(300A,0015)	Nominal Beam Energy Unit	CS	1
(300A,0016)	Dose Reference Description	LO	1
(300A,0018)	Dose Reference Point Coordinates	DS	3
(300A,001A)	Nominal Prior Dose	DS	1
(300A,0020)	Dose Reference Type	CS	1
(300A,0021)	Constraint Weight	DS	1
(300A,0022)	Delivery Warning Dose	DS	1
(300A,0023)	Delivery Maximum Dose	DS	1
(300A,0025)	Target Minimum Dose	DS	1
(300A,0026)	Target Prescription Dose	DS	1

(300A,0027)	Target Maximum Dose	DS	1
(300A,0028)	Target Underdose Volume Fraction	DS	1
(300A,002A)	Organ at Risk Full-volume Dose	DS	1
(300A,002B)	Organ at Risk Limit Dose	DS	1
(300A,002C)	Organ at Risk Maximum Dose	DS	1
(300A,002D)	Organ at Risk Overdose Volume Fraction	DS	1
(300A,0040)	Tolerance Table Sequence	SQ	1
(300A,0042)	Tolerance Table Number	IS	1
(300A,0043)	Tolerance Table Label	SH	1
(300A,0044)	Gantry Angle Tolerance	DS	1
(300A,0046)	Beam Limiting Device Angle Tolerance	DS	1
(300A,0048)	Beam Limiting Device Tolerance Sequence	SQ	1
(300A,004A)	Beam Limiting Device Position Tolerance	DS	1
(300A,004B)	Snout Position Tolerance	FL	1
(300A,004C)	Patient Support Angle Tolerance	DS	1
(300A,004E)	Table Top Eccentric Angle Tolerance	DS	1
(300A,004F)	Table Top Pitch Angle Tolerance	FL	1
(300A,0050)	Table Top Roll Angle Tolerance	FL	1
(300A,0051)	Table Top Vertical Position Tolerance	DS	1
(300A,0052)	Table Top Longitudinal Position Tolerance	DS	1
(300A,0053)	Table Top Lateral Position Tolerance	DS	1
(300A,0055)	RT Plan Relationship	CS	1
(300A,0070)	Fraction Group Sequence	SQ	1
(300A,0071)	Fraction Group Number	IS	1
(300A,0072)	Fraction Group Description	LO	1
(300A,0078)	Number of Fractions Planned	IS	1
(300A,0079)	Number of Fraction Pattern Digits Per Day	IS	1
(300A,007A)	Repeat Fraction Cycle Length	IS	1
(300A,007B)	Fraction Pattern	LT	1
(300A,0080)	Number of Beams	IS	1
(300A,0082)	Beam Dose Specification Point	DS	3
(300A,0084)	Beam Dose	DS	1
(300A,0086)	Beam Meterset	DS	1
(300A,0088)	Beam Dose Point Depth	FL	1

(300A,0089)	Beam Dose Point Equivalent Depth	FL	1
(300A,008A)	Beam Dose Point SSD	FL	1
(300A,00A0)	Number of Brachy Application Setups	IS	1
(300A,00A2)	Brachy Application Setup Dose Specification Point	DS	3
(300A,00A4)	Brachy Application Setup Dose	DS	1
(300A,00B0)	Beam Sequence	SQ	1
(300A,00B2)	Treatment Machine Name	SH	1
(300A,00B3)	Primary Dosimeter Unit	CS	1
(300A,00B4)	Source-Axis Distance	DS	1
(300A,00B6)	Beam Limiting Device Sequence	SQ	1
(300A,00B8)	RT Beam Limiting Device Type	CS	1
(300A,00BA)	Source to Beam Limiting Device Distance	DS	1
(300A,00BB)	Isocenter to Beam Limiting Device Distance	FL	1
(300A,00BC)	Number of Leaf/Jaw Pairs	IS	1
(300A,00BE)	Leaf Position Boundaries	DS	3-n
(300A,00C0)	Beam Number	IS	1
(300A,00C2)	Beam Name	LO	1
(300A,00C3)	Beam Description	ST	1
(300A,00C4)	Beam Type	CS	1
(300A,00C6)	Radiation Type	CS	1
(300A,00C7)	High-Dose Technique Type	CS	1
(300A,00C8)	Reference Image Number	IS	1
(300A,00CA)	Planned Verification Image Sequence	SQ	1
(300A,00CC)	Imaging Device-Specific Acquisition Parameters	LO	1-n
(300A,00CE)	Treatment Delivery Type	CS	1
(300A,00D0)	Number of Wedges	IS	1
(300A,00D1)	Wedge Sequence	SQ	1
(300A,00D2)	Wedge Number	IS	1
(300A,00D3)	Wedge Type	CS	1
(300A,00D4)	Wedge ID	SH	1
(300A,00D5)	Wedge Angle	IS	1
(300A,00D6)	Wedge Factor	DS	1
(300A,00D7)	Total Wedge Tray Water-Equivalent Thickness	FL	1
(300A,00D8)	Wedge Orientation	DS	1

(300A,00D9)	Isocenter to Wedge Tray Distance	FL	1
(300A,00DA)	Source to Wedge Tray Distance	DS	1
(300A,00DB)	Wedge Thin Edge Position	FL	1
(300A,00DC)	Bolus ID	SH	1
(300A,00DD)	Bolus Description	ST	1
(300A,00E0)	Number of Compensators	IS	1
(300A,00E1)	Material ID	SH	1
(300A,00E2)	Total Compensator Tray Factor	DS	1
(300A,00E3)	Compensator Sequence	SQ	1
(300A,00E4)	Compensator Number	IS	1
(300A,00E5)	Compensator ID	SH	1
(300A,00E6)	Source to Compensator Tray Distance	DS	1
(300A,00E7)	Compensator Rows	IS	1
(300A,00E8)	Compensator Columns	IS	1
(300A,00E9)	Compensator Pixel Spacing	DS	2
(300A,00EA)	Compensator Position	DS	2
(300A,00EB)	Compensator Transmission Data	DS	1-n
(300A,00EC)	Compensator Thickness Data	DS	1-n
(300A,00ED)	Number of Boli	IS	1
(300A,00EE)	Compensator Type	CS	1
(300A,00F0)	Number of Blocks	IS	1
(300A,00F2)	Total Block Tray Factor	DS	1
(300A,00F3)	Total Block Tray Water-Equivalent Thickness	FL	1
(300A,00F4)	Block Sequence	SQ	1
(300A,00F5)	Block Tray ID	SH	1
(300A,00F6)	Source to Block Tray Distance	DS	1
(300A,00F7)	Isocenter to Block Tray Distance	FL	1
(300A,00F8)	Block Type	CS	1
(300A,00F9)	Accessory Code	LO	1
(300A,00FA)	Block Divergence	CS	1
(300A,00FB)	Block Mounting Position	CS	1
(300A,00FC)	Block Number	IS	1
(300A,00FE)	Block Name	LO	1
(300A,0100)	Block Thickness	DS	1

(300A,0102)	Block Transmission	DS	1
(300A,0104)	Block Number of Points	IS	1
(300A,0106)	Block Data	DS	2-2n
(300A,0107)	Applicator Sequence	SQ	1
(300A,0108)	Applicator ID	SH	1
(300A,0109)	Applicator Type	CS	1
(300A,010A)	Applicator Description	LO	1
(300A,010C)	Cumulative Dose Reference Coefficient	DS	1
(300A,010E)	Final Cumulative Meterset Weight	DS	1
(300A,0110)	Number of Control Points	IS	1
(300A,0111)	Control Point Sequence	SQ	1
(300A,0112)	Control Point Index	IS	1
(300A,0114)	Nominal Beam Energy	DS	1
(300A,0115)	Dose Rate Set	DS	1
(300A,0116)	Wedge Position Sequence	SQ	1
(300A,0118)	Wedge Position	CS	1
(300A,011A)	Beam Limiting Device Position Sequence	SQ	1
(300A,011C)	Leaf/Jaw Positions	DS	2-2n
(300A,011E)	Gantry Angle	DS	1
(300A,011F)	Gantry Rotation Direction	CS	1
(300A,0120)	Beam Limiting Device Angle	DS	1
(300A,0121)	Beam Limiting Device Rotation Direction	CS	1
(300A,0122)	Patient Support Angle	DS	1
(300A,0123)	Patient Support Rotation Direction	CS	1
(300A,0124)	Table Top Eccentric Axis Distance	DS	1
(300A,0125)	Table Top Eccentric Angle	DS	1
(300A,0126)	Table Top Eccentric Rotation Direction	CS	1
(300A,0128)	Table Top Vertical Position	DS	1
(300A,0129)	Table Top Longitudinal Position	DS	1
(300A,012A)	Table Top Lateral Position	DS	1
(300A,012C)	Isocenter Position	DS	3
(300A,012E)	Surface Entry Point	DS	3
(300A,0130)	Source to Surface Distance	DS	1
(300A,0134)	Cumulative Meterset Weight	DS	1



(300A,0140)	Table Top Pitch Angle	FL	1
(300A,0142)	Table Top Pitch Rotation Direction	CS	1
(300A,0144)	Table Top Roll Angle	FL	1
(300A,0146)	Table Top Roll Rotation Direction	CS	1
(300A,0148)	Head Fixation Angle	FL	1
(300A,014A)	Gantry Pitch Angle	FL	1
(300A,014C)	Gantry Pitch Rotation Direction	CS	1
(300A,014E)	Gantry Pitch Angle Tolerance	FL	1
(300A,0180)	Patient Setup Sequence	SQ	1
(300A,0182)	Patient Setup Number	IS	1
(300A,0183)	Patient Setup Label	LO	1
(300A,0184)	Patient Additional Position	LO	1
(300A,0190)	Fixation Device Sequence	SQ	1
(300A,0192)	Fixation Device Type	CS	1
(300A,0194)	Fixation Device Label	SH	1
(300A,0196)	Fixation Device Description	ST	1
(300A,0198)	Fixation Device Position	SH	1
(300A,0199)	Fixation Device Pitch Angle	FL	1
(300A,019A)	Fixation Device Roll Angle	FL	1
(300A,01A0)	Shielding Device Sequence	SQ	1
(300A,01A2)	Shielding Device Type	CS	1
(300A,01A4)	Shielding Device Label	SH	1
(300A,01A6)	Shielding Device Description	ST	1
(300A,01A8)	Shielding Device Position	SH	1
(300A,01B0)	Setup Technique	CS	1
(300A,01B2)	Setup Technique Description	ST	1
(300A,01B4)	Setup Device Sequence	SQ	1
(300A,01B6)	Setup Device Type	CS	1
(300A,01B8)	Setup Device Label	SH	1
(300A,01BA)	Setup Device Description	ST	1
(300A,01BC)	Setup Device Parameter	DS	1
(300A,01D0)	Setup Reference Description	ST	1
(300A,01D2)	Table Top Vertical Setup Displacement	DS	1
(300A,01D4)	Table Top Longitudinal Setup Displacement	DS	1

(300A,01D6)	Table Top Lateral Setup Displacement	DS	1
(300A,0200)	Brachy Treatment Technique	CS	1
(300A,0202)	Brachy Treatment Type	CS	1
(300A,0206)	Treatment Machine Sequence	SQ	1
(300A,0210)	Source Sequence	SQ	1
(300A,0212)	Source Number	IS	1
(300A,0214)	Source Type	CS	1
(300A,0216)	Source Manufacturer	LO	1
(300A,0218)	Active Source Diameter	DS	1
(300A,021A)	Active Source Length	DS	1
(300A,0222)	Source Encapsulation Nominal Thickness	DS	1
(300A,0224)	Source Encapsulation Nominal Transmission	DS	1
(300A,0226)	Source Isotope Name	LO	1
(300A,0228)	Source Isotope Half Life	DS	1
(300A,0229)	Source Strength Units	CS	1
(300A,022A)	Reference Air Kerma Rate	DS	1
(300A,022B)	Source Strength	DS	1
(300A,022C)	Source Strength Reference Date	DA	1
(300A,022E)	Source Strength Reference Time	TM	1
(300A,0230)	Application Setup Sequence	SQ	1
(300A,0232)	Application Setup Type	CS	1
(300A,0234)	Application Setup Number	IS	1
(300A,0236)	Application Setup Name	LO	1
(300A,0238)	Application Setup Manufacturer	LO	1
(300A,0240)	Template Number	IS	1
(300A,0242)	Template Type	SH	1
(300A,0244)	Template Name	LO	1
(300A,0250)	Total Reference Air Kerma	DS	1
(300A,0260)	Brachy Accessory Device Sequence	SQ	1
(300A,0262)	Brachy Accessory Device Number	IS	1
(300A,0263)	Brachy Accessory Device ID	SH	1
(300A,0264)	Brachy Accessory Device Type	CS	1
(300A,0266)	Brachy Accessory Device Name	LO	1
(300A,026A)	Brachy Accessory Device Nominal Thickness	DS	1

(300A,026C)	Brachy Accessory Device Nominal Transmission	DS	1
(300A,0280)	Channel Sequence	SQ	1
(300A,0282)	Channel Number	IS	1
(300A,0284)	Channel Length	DS	1
(300A,0286)	Channel Total Time	DS	1
(300A,0288)	Source Movement Type	CS	1
(300A,028A)	Number of Pulses	IS	1
(300A,028C)	Pulse Repetition Interval	DS	1
(300A,0290)	Source Applicator Number	IS	1
(300A,0291)	Source Applicator ID	SH	1
(300A,0292)	Source Applicator Type	CS	1
(300A,0294)	Source Applicator Name	LO	1
(300A,0296)	Source Applicator Length	DS	1
(300A,0298)	Source Applicator Manufacturer	LO	1
(300A,029C)	Source Applicator Wall Nominal Thickness	DS	1
(300A,029E)	Source Applicator Wall Nominal Transmission	DS	1
(300A,02A0)	Source Applicator Step Size	DS	1
(300A,02A2)	Transfer Tube Number	IS	1
(300A,02A4)	Transfer Tube Length	DS	1
(300A,02B0)	Channel Shield Sequence	SQ	1
(300A,02B2)	Channel Shield Number	IS	1
(300A,02B3)	Channel Shield ID	SH	1
(300A,02B4)	Channel Shield Name	LO	1
(300A,02B8)	Channel Shield Nominal Thickness	DS	1
(300A,02BA)	Channel Shield Nominal Transmission	DS	1
(300A,02C8)	Final Cumulative Time Weight	DS	1
(300A,02D0)	Brachy Control Point Sequence	SQ	1
(300A,02D2)	Control Point Relative Position	DS	1
(300A,02D4)	Control Point 3D Position	DS	3
(300A,02D6)	Cumulative Time Weight	DS	1
(300A,02E0)	Compensator Divergence	CS	1
(300A,02E1)	Compensator Mounting Position	CS	1
(300A,02E2)	Source to Compensator Distance	DS	1-n
(300A,02E3)	Total Compensator Tray Water-Equivalent Thickness	FL	1

(300A,02E4)	Isocenter to Compensator Tray Distance	FL	1
(300A,02E5)	Compensator Column Offset	FL	1
(300A,02E6)	Isocenter to Compensator Distances	FL	1-n
(300A,02E7)	Compensator Relative Stopping Power Ratio	FL	1
(300A,02E8)	Compensator Milling Tool Diameter	FL	1
(300A,02EA)	Ion Range Compensator Sequence	SQ	1
(300A,0302)	Radiation Mass Number	IS	1
(300A,0304)	Radiation Atomic Number	IS	1
(300A,0306)	Radiation Charge State	SS	1
(300A,0308)	Scan Mode	CS	1
(300A,030A)	Virtual Source-Axis Distances	FL	2
(300A,030C)	Snout Sequence	SQ	1
(300A,030D)	Snout Position	FL	1
(300A,030F)	Snout ID	SH	1
(300A,0312)	Number of Range Shifters	IS	1
(300A,0314)	Range Shifter Sequence	SQ	1
(300A,0316)	Range Shifter Number	IS	1
(300A,0318)	Range Shifter ID	SH	1
(300A,0320)	Range Shifter Type	CS	1
(300A,0322)	Range Shifter Description	LO	1
(300A,0330)	Number of Lateral Spreading Devices	IS	1
(300A,0332)	Lateral Spreading Device Sequence	SQ	1
(300A,0334)	Lateral Spreading Device Number	IS	1
(300A,0336)	Lateral Spreading Device ID	SH	1
(300A,0338)	Lateral Spreading Device Type	CS	1
(300A,033A)	Lateral Spreading Device Description	LO	1
(300A,033C)	Lateral Spreading Device Water Equivalent Thickness	FL	1
(300A,0340)	Number of Range Modulators	IS	1
(300A,0342)	Range Modulator Sequence	SQ	1
(300A,0344)	Range Modulator Number	IS	1

(300A,0346)	Range Modulator ID	SH	1
(300A,0348)	Range Modulator Type	CS	1
(300A,034A)	Range Modulator Description	LO	1
(300A,034C)	Beam Current Modulation ID	SH	1
(300A,0350)	Patient Support Type	CS	1
(300A,0352)	Patient Support ID	SH	1
(300A,0354)	Patient Support Accessory Code	LO	1
(300A,0356)	Fixation Light Azimuthal Angle	FL	1
(300A,0358)	Fixation Light Polar Angle	FL	1
(300A,035A)	Meterset Rate	FL	1
(300A,0360)	Range Shifter Settings Sequence	SQ	1
(300A,0362)	Range Shifter Setting	LO	1
(300A,0364)	Isocenter to Range Shifter Distance	FL	1
(300A,0366)	Range Shifter Water Equivalent Thickness	FL	1
(300A,0370)	Lateral Spreading Device Settings Sequence	SQ	1
(300A,0372)	Lateral Spreading Device Setting	LO	1
(300A,0374)	Isocenter to Lateral Spreading Device Distance	FL	1
(300A,0380)	Range Modulator Settings Sequence	SQ	1
(300A,0382)	Range Modulator Gating Start Value	FL	1
(300A,0384)	Range Modulator Gating Stop Value	FL	1
(300A,0386)	Range Modulator Gating Start Water Equivalent Thickness	FL	1
(300A,0388)	Range Modulator Gating Stop Water Equivalent Thickness	FL	1
(300A,038A)	Isocenter to Range Modulator Distance	FL	1
(300A,0390)	Scan Spot Tune ID	SH	1
(300A,0392)	Number of Scan Spot Positions	IS	1
(300A,0394)	Scan Spot Position Map	FL	1-n
(300A,0396)	Scan Spot Meterset Weights	FL	1-n
(300A,0398)	Scanning Spot Size	FL	2
(300A,039A)	Number of Paintings	IS	1

(300A,03A0)	Ion Tolerance Table Sequence	SQ	1
(300A,03A2)	Ion Beam Sequence	SQ	1
(300A,03A4)	Ion Beam Limiting Device Sequence	SQ	1
(300A,03A6)	Ion Block Sequence	SQ	1
(300A,03A8)	Ion Control Point Sequence	SQ	1
(300A,03AA)	Ion Wedge Sequence	SQ	1
(300A,03AC)	Ion Wedge Position Sequence	SQ	1
(300A,0401)	Referenced Setup Image Sequence	SQ	1
(300A,0402)	Setup Image Comment	ST	1
(300A,0410)	Motion Synchronization Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(300C,0002)	Referenced RT Plan Sequence	SQ	1
(300C,0004)	Referenced Beam Sequence	SQ	1
(300C,0006)	Referenced Beam Number	IS	1
(300C,0007)	Referenced Reference Image Number	IS	1
(300C,0008)	Start Cumulative Meterset Weight	DS	1
(300C,0009)	End Cumulative Meterset Weight	DS	1
(300C,000A)	Referenced Brachy Application Setup Sequence	SQ	1
(300C,000C)	Referenced Brachy Application Setup Number	IS	1
(300C,000E)	Referenced Source Number	IS	1
(300C,0020)	Referenced Fraction Group Sequence	SQ	1
(300C,0022)	Referenced Fraction Group Number	IS	1
(300C,0040)	Referenced Verification Image Sequence	SQ	1
(300C,0042)	Referenced Reference Image Sequence	SQ	1
(300C,0050)	Referenced Dose Reference Sequence	SQ	1
(300C,0051)	Referenced Dose Reference Number	IS	1
(300C,0055)	Brachy Referenced Dose Reference Sequence	SQ	1
(300C,0060)	Referenced Structure Set Sequence	SQ	1
(300C,006A)	Referenced Patient Setup Number	IS	1
(300C,0080)	Referenced Dose Sequence	SQ	1
(300C,00A0)	Referenced Tolerance Table Number	IS	1

(300C,00B0)	Referenced Bolus Sequence	SQ	1
(300C,00C0)	Referenced Wedge Number	IS	1
(300C,00D0)	Referenced Compensator Number	IS	1
(300C,00E0)	Referenced Block Number	IS	1
(300C,00F0)	Referenced Control Point Index	IS	1
(300C,00F2)	Referenced Control Point Sequence	SQ	1
(300C,00F4)	Referenced Start Control Point Index	IS	1
(300C,00F6)	Referenced Stop Control Point Index	IS	1
(300C,0100)	Referenced Range Shifter Number	IS	1
(300C,0102)	Referenced Lateral Spreading Device Number	IS	1
(300C,0104)	Referenced Range Modulator Number	IS	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(300E,0002)	Approval Status	CS	1
(300E,0004)	Review Date	DA	1
(300E,0005)	Review Time	TM	1
(300E,0008)	Reviewer Name	PN	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
<i>(4000,0010)</i>	<i>Arbitrary</i>	<i>LT</i>	<i>1 RET</i>
<i>(4000,4000)</i>	<i>Text Comments</i>	<i>LT</i>	<i>1 RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
<i>(4008,0040)</i>	<i>Results ID</i>	<i>SH</i>	<i>1 RET</i>
<i>(4008,0042)</i>	<i>Results ID Issuer</i>	<i>LO</i>	<i>1 RET</i>
<i>(4008,0050)</i>	<i>Referenced Interpretation Sequence</i>	<i>SQ</i>	<i>1 RET</i>
<i>(4008,0100)</i>	<i>Interpretation Recorded Date</i>	<i>DA</i>	<i>1 RET</i>
<i>(4008,0101)</i>	<i>Interpretation Recorded Time</i>	<i>TM</i>	<i>1 RET</i>
<i>(4008,0102)</i>	<i>Interpretation Recorder</i>	<i>PN</i>	<i>1 RET</i>
<i>(4008,0103)</i>	<i>Reference to Recorded Sound</i>	<i>LO</i>	<i>1 RET</i>
<i>(4008,0108)</i>	<i>Interpretation Transcription Date</i>	<i>DA</i>	<i>1 RET</i>
<i>(4008,0109)</i>	<i>Interpretation Transcription Time</i>	<i>TM</i>	<i>1 RET</i>

(4008,010A)	<i>Interpretation Transcriber</i>	PN	1	RET
(4008,010B)	<i>Interpretation Text</i>	ST	1	RET
(4008,010C)	<i>Interpretation Author</i>	PN	1	RET
(4008,0111)	<i>Interpretation Approver Sequence</i>	SQ	1	RET
(4008,0112)	<i>Interpretation Approval Date</i>	DA	1	RET
(4008,0113)	<i>Interpretation Approval Time</i>	TM	1	RET
(4008,0114)	<i>Physician Approving Interpretation</i>	PN	1	RET
(4008,0115)	<i>Interpretation Diagnosis Description</i>	LT	1	RET
(4008,0117)	<i>Interpretation Diagnosis Code Sequence</i>	SQ	1	RET
(4008,0118)	<i>Results Distribution List Sequence</i>	SQ	1	RET
(4008,0119)	<i>Distribution Name</i>	PN	1	RET
(4008,011A)	<i>Distribution Address</i>	LO	1	RET
(4008,0200)	<i>Interpretation ID</i>	SH	1	RET
(4008,0202)	<i>Interpretation ID Issuer</i>	LO	1	RET
(4008,0210)	<i>Interpretation Type ID</i>	CS	1	RET
(4008,0212)	<i>Interpretation Status ID</i>	CS	1	RET
(4008,0300)	<i>Impressions</i>	ST	1	RET
(4008,4000)	<i>Results Comments</i>	ST	1	RET

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(4FFE,0001)	MAC Parameters Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(50xx,0005)	<i>Curve Dimensions</i>	US	1 RET
(50xx,0010)	<i>Number of Points</i>	US	1 RET
(50xx,0020)	<i>Type of Data</i>	CS	1 RET
(50xx,0022)	<i>Curve Description</i>	LO	1 RET
(50xx,0030)	<i>Axis Units</i>	SH	1-n RET
(50xx,0040)	<i>Axis Labels</i>	SH	1-n RET
(50xx,0103)	<i>Data Value Representation</i>	US	1 RET
(50xx,0104)	<i>Minimum Coordinate Value</i>	US	1-n RET
(50xx,0105)	<i>Maximum Coordinate Value</i>	US	1-n RET
(50xx,0106)	<i>Curve Range</i>	SH	1-n RET



(50xx,0110)	Curve Data Descriptor	US	1-n	RET
(50xx,0112)	Coordinate Start Value	US	1-n	RET
(50xx,0114)	Coordinate Step Value	US	1-n	RET
(50xx,1001)	Curve Activation Layer	CS	1	RET
(50xx,2000)	Audio Type	US	1	RET
(50xx,2002)	Audio Sample Format	US	1	RET
(50xx,2004)	Number of Channels	US	1	RET
(50xx,2006)	Number of Samples	UL	1	RET
(50xx,2008)	Sample Rate	UL	1	RET
(50xx,200A)	Total Time	UL	1	RET
(50xx,200C)	Audio Sample Data	OW or OB	1	RET
(50xx,200E)	Audio Comments	LT	1	RET
(50xx,2500)	Curve Label	LO	1	RET
(50xx,2600)	Curve Referenced Overlay Sequence	SQ	1	RET
(50xx,2610)	Curve Referenced Overlay Group	US	1	RET
(50xx,3000)	Curve Data	OW or OB	1	RET

Tag	Name	VR	VM
(5200,9229)	Shared Functional Groups Sequence	SQ	1
(5200,9230)	Per-frame Functional Groups Sequence	SQ	1

Tag	Name	VR	VM
(5400,0100)	Waveform Sequence	SQ	1
(5400,0110)	Channel Minimum Value	OB or OW	1
(5400,0112)	Channel Maximum Value	OB or OW	1
(5400,1004)	Waveform Bits Allocated	US	1
(5400,1006)	Waveform Sample Interpretation	CS	1
(5400,100A)	Waveform Padding Value	OB or OW	1
(5400,1010)	Waveform Data	OB or OW	1

Tag	Name	VR	VM
(5600,0010)	First Order Phase Correction Angle	OF	1
(5600,0020)	Spectroscopy Data	OF	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>	
(60xx,0010)	Overlay Rows	US	1	
(60xx,0011)	Overlay Columns	US	1	
(60xx,0012)	Overlay Planes	US	1	
(60xx,0015)	Number of Frames in Overlay	IS	1	
(60xx,0022)	Overlay Description	LO	1	
(60xx,0040)	Overlay Type	CS	1	
(60xx,0045)	Overlay Subtype	LO	1	
(60xx,0050)	Overlay Origin	SS	2	
(60xx,0051)	Image Frame Origin	US	1	
(60xx,0052)	Overlay Plane Origin	US	1	
<i>(60xx,0060)</i>	<i>Overlay Compression Code</i>	CS	1	<i>RET</i>
(60xx,0100)	Overlay Bits Allocated	US	1	
(60xx,0102)	Overlay Bit Position	US	1	
<i>(60xx,0110)</i>	<i>Overlay Format</i>	CS	1	<i>RET</i>
<i>(60xx,0200)</i>	<i>Overlay Location</i>	US	1	<i>RET</i>
(60xx,1001)	Overlay Activation Layer	CS	1	
(60xx,1100)	Overlay Descriptor – Gray	US	1	<i>RET</i>
(60xx,1101)	Overlay Descriptor – Red	US	1	<i>RET</i>
(60xx,1102)	Overlay Descriptor – Green	US	1	<i>RET</i>
(60xx,1103)	Overlay Descriptor – Blue	US	1	<i>RET</i>
(60xx,1200)	Overlays- Gray	US	1-n	<i>RET</i>
(60xx,1201)	Overlays – Red	US	1-n	<i>RET</i>
(60xx,1202)	Overlays – Green	US	1-n	<i>RET</i>
(60xx,1203)	Overlays- Blue	US	1-n	<i>RET</i>
(60xx,1301)	ROI Area	IS	1	
(60xx,1302)	ROI Mean	DS	1	
(60xx,1303)	ROI Standard Deviation	DS	1	
(60xx,1500)	Overlay Label	LO	1	
(60xx,3000)	Overlay Data	OB or OW	1	
<i>(60xx,4000)</i>	<i>Overlay Comments</i>	<i>LT</i>	<i>1</i>	<i>RET</i>

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(7FE0,0010)	Pixel Data	OW or OB	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(FFFA,FFFA)	Digital Signatures Sequence	SQ	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(FFFC,FFFC)	Data Set Trailing Padding	OB	1

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(FFFE,E000)	Item	see note	1
(FFFE,E00D)	Item Delimitation Item	see note	1
(FFFE,E0DD)	Sequence Delimitation Item	see note	1

Note: The VR for Data Elements, Item (FFFE,E000), Item Delimitation Item (FFFE,E00D), and Sequence Delimitation Item (FFFE,E0DD) do not exist. See PS 3.5 for explanation.

## 7 Registry of DICOM File Meta Elements

This section specifies the File Meta Elements needed to support the formatting of the File Meta Information of the DICOM File Format (See PS 3.10).

<b>Tag</b>	<b>Name</b>	<b>VR</b>	<b>VM</b>
(0002,0000)	Group Length	UL	1
(0002,0001)	File Meta Information Version	OB	1
(0002,0002)	Media Storage SOP Class UID	UI	1
(0002,0003)	Media Storage SOP Instance UID	UI	1
(0002,0010)	Transfer Syntax UID	UI	1
(0002,0012)	Implementation Class UID	UI	1
(0002,0013)	Implementation Version Name	SH	1
(0002,0016)	Source Application Entity Title	AE	1
(0002,0100)	Private Information Creator UID	UI	1
(0002,0102)	Private Information	OB	1

## 8 Registry of DICOM directory structuring elements

Tag	Name	VR	VM	
(0004,0000)	Group Length	UL	1	
(0004,1130)	File-set ID	CS	1	
(0004,1141)	File-set Descriptor File ID	CS	1-8	
(0004,1142)	Specific Character Set of File-set Descriptor File	CS	1	
(0004,1200)	Offset of the First Directory Record of the Root Directory Entity	UL	1	
(0004,1202)	Offset of the Last Directory Record of the Root Directory Entity	UL	1	
(0004,1212)	File-set Consistency Flag	US	1	
(0004,1220)	Directory Record Sequence	SQ	1	
(0004,1400)	Offset of the Next Directory Record	UL	1	
(0004,1410)	Record In-use Flag	US	1	
(0004,1420)	Offset of Referenced Lower-Level Directory Entity	UL	1	
(0004,1430)	Directory Record Type	CS	1	
(0004,1432)	Private Record UID	UI	1	
(0004,1500)	Referenced File ID	CS	1-8	
(0004,1504)	<i>MRDR Directory Record Offset</i>	<i>UL</i>	<i>1</i>	<i>RET</i>
(0004,1510)	Referenced SOP Class UID in File	UI	1	
(0004,1511)	Referenced SOP Instance UID in File	UI	1	
(0004,1512)	Referenced Transfer Syntax UID in File	UI	1	
(0004,151A)	Referenced Related General SOP Class UID in File	UI	1-n	
(0004,1600)	<i>Number of References</i>	<i>UL</i>	<i>1</i>	<i>RET</i>

## Annex A Registry of DICOM unique identifiers (UID) (Normative)

Table A-1 lists the UID values that are registered and used throughout the Parts of the DICOM Standard. This central registry ensures that when additional UIDs are assigned, non-duplicate values are assigned.

**Table A-1  
UID VALUES**

UID Value	UID NAME	UID TYPE	Part
1.2.840.10008.1.1	Verification SOP Class	SOP Class	PS 3.4
1.2.840.10008.1.2	Implicit VR Little Endian: Default Transfer Syntax for DICOM	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.1	Explicit VR Little Endian	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.1.99	Deflated Explicit VR Little Endian	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.2	Explicit VR Big Endian	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.50	JPEG Baseline (Process 1): Default Transfer Syntax for Lossy JPEG 8 Bit Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.51	JPEG Extended (Process 2 & 4): Default Transfer Syntax for Lossy JPEG 12 Bit Image Compression (Process 4 only)	Transfer Syntax	PS 3.5
<i>1.2.840.10008.1.2.4.52</i>	<i>JPEG Extended (Process 3 &amp; 5) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>
<i>1.2.840.10008.1.2.4.53</i>	<i>JPEG Spectral Selection, Non- Hierarchical (Process 6 &amp; 8) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>
<i>1.2.840.10008.1.2.4.54</i>	<i>JPEG Spectral Selection, Non- Hierarchical (Process 7 &amp; 9) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>
<i>1.2.840.10008.1.2.4.55</i>	<i>JPEG Full Progression, Non- Hierarchical (Process 10 &amp; 12) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>
<i>1.2.840.10008.1.2.4.56</i>	<i>JPEG Full Progression, Non- Hierarchical (Process 11 &amp; 13) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>
1.2.840.10008.1.2.4.57	JPEG Lossless, Non-Hierarchical (Process 14)	Transfer Syntax	PS 3.5
<i>1.2.840.10008.1.2.4.58</i>	<i>JPEG Lossless, Non-Hierarchical (Process 15) (Retired)</i>	<i>Transfer Syntax</i>	<i>PS 3.5</i>

1.2.840.10008.1.2.4.59	JPEG Extended, Hierarchical (Process 16 & 18) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.60	JPEG Extended, Hierarchical (Process 17 & 19) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.61	JPEG Spectral Selection, Hierarchical (Process 20 & 22) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.62	JPEG Spectral Selection, Hierarchical (Process 21 & 23) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.63	JPEG Full Progression, Hierarchical (Process 24 & 26) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.64	JPEG Full Progression, Hierarchical (Process 25 & 27) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.65	JPEG Lossless, Hierarchical (Process 28) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.66	JPEG Lossless, Hierarchical (Process 29) (Retired)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.70	JPEG Lossless, Non-Hierarchical, First-Order Prediction (Process 14 [Selection Value 1]): Default Transfer Syntax for Lossless JPEG Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.80	JPEG-LS Lossless Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.81	JPEG-LS Lossy (Near-Lossless) Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.90	JPEG 2000 Image Compression (Lossless Only)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.91	JPEG 2000 Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.92	JPEG 2000 Part 2 Multi-component Image Compression (Lossless Only)	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.93	JPEG 2000 Part 2 Multi-component Image Compression	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.94	JPIP Referenced	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.95	JPIP Referenced Deflate	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.4.100	MPEG2 Main Profile @ Main Level	Transfer Syntax	PS 3.5

1.2.840.10008.1.2.5	RLE Lossless	Transfer Syntax	PS 3.5
1.2.840.10008.1.2.6.1	RFC 2557 MIME encapsulation	Transfer Syntax	PS 3.10
1.2.840.10008.1.3.10	Media Storage Directory Storage	SOP Class	PS 3.4
1.2.840.10008.1.4.1.1	Talairach Brain Atlas Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.2	SPM2 T1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.3	SPM2 T2 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.4	SPM2 PD Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.5	SPM2 EPI Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.6	SPM2 FIL T1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.7	SPM2 PET Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.8	SPM2 TRANSM Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.9	SPM2 SPECT Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.10	SPM2 GRAY Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.11	SPM2 WHITE Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.12	SPM2 CSF Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.13	SPM2 BRAINMASK Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.14	SPM2 AVG305T1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.15	SPM2 AVG152T1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.16	SPM2 AVG152T2 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.1.17	SPM2 AVG152PD Frame of Reference	Well-known frame of reference	



1.2.840.10008.1.4.1.18	SPM2 SINGLESUBJT1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.2.1	ICBM 452 T1 Frame of Reference	Well-known frame of reference	
1.2.840.10008.1.4.2.2	ICBM Single Subject MRI Frame of Reference	Well-known frame of reference	
<i>1.2.840.10008.1.9</i>	<i>Basic Study Content Notification SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.1.20.1	Storage Commitment Push Model SOP Class	SOP Class	PS 3.4
1.2.840.10008.1.20.1.1	Storage Commitment Push Model SOP Instance	Well-known SOP Instance	PS 3.4
<i>1.2.840.10008.1.20.2</i>	<i>Storage Commitment Pull Model SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.1.20.2.1</i>	<i>Storage Commitment Pull Model SOP Instance (Retired)</i>	<i>Well-known SOP Instance</i>	<i>PS 3.4</i>
1.2.840.10008.1.40	Procedural Event Logging SOP Class	SOP Class	PS 3.4
1.2.840.10008.1.40.1	Procedural Event Logging SOP Instance	Well-known SOP Instance	PS 3.4
1.2.840.10008.2.6.1	DICOM UID Registry	DICOM UIDs as a Coding Scheme	PS 3.6
1.2.840.10008.2.16.4	DICOM Controlled Terminology	Coding Scheme	PS 3.16
1.2.840.10008.3.1.1.1	DICOM Application Context Name	Application Context Name	PS 3.7
<i>1.2.840.10008.3.1.2.1.1</i>	<i>Detached Patient Management SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.3.1.2.1.4</i>	<i>Detached Patient Management Meta SOP Class (Retired)</i>	<i>Meta SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.3.1.2.2.1</i>	<i>Detached Visit Management SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.3.1.2.3.1</i>	<i>Detached Study Management SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.3.1.2.3.2</i>	<i>Study Component Management SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.3.1.2.3.3	Modality Performed Procedure Step SOP Class	SOP Class	PS 3.4
1.2.840.10008.3.1.2.3.4	Modality Performed Procedure Step Retrieve SOP Class	SOP Class	PS 3.4
1.2.840.10008.3.1.2.3.5	Modality Performed Procedure	SOP Class	PS 3.4

	Step Notification SOP Class		
1.2.840.10008.3.1.2.5.1	<i>Detached Results Management SOP Class (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.3.1.2.5.4	<i>Detached Results Management Meta SOP Class (Retired)</i>	Meta SOP Class	PS 3.4
1.2.840.10008.3.1.2.5.5	<i>Detached Study Management Meta SOP Class (Retired)</i>	Meta SOP Class	PS 3.4
1.2.840.10008.3.1.2.6.1	<i>Detached Interpretation Management SOP Class (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.4.2	Storage Service Class	Service Class	PS 3.4
1.2.840.10008.5.1.1.1	Basic Film Session SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.2	Basic Film Box SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.4	Basic Grayscale Image Box SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.4.1	Basic Color Image Box SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.4.2	<i>Referenced Image Box SOP Class (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.5.1.1.9	Basic Grayscale Print Management Meta SOP Class	Meta SOP Class	PS 3.4
1.2.840.10008.5.1.1.9.1	<i>Referenced Grayscale Print Management Meta SOP Class (Retired)</i>	Meta SOP Class	PS 3.4
1.2.840.10008.5.1.1.14	Print Job SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.15	Basic Annotation Box SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.16	Printer SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.16.376	Printer Configuration Retrieval SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.17	Printer SOP Instance	Well-known Printer SOP Instance	PS 3.4
1.2.840.10008.5.1.1.17.376	Printer Configuration Retrieval SOP Instance	Well-known Printer SOP Instance	PS 3.4
1.2.840.10008.5.1.1.18	Basic Color Print Management Meta SOP Class	Meta SOP Class	PS 3.4
1.2.840.10008.5.1.1.18.1	<i>Referenced Color Print Management Meta SOP Class (Retired)</i>	Meta SOP Class	PS 3.4
1.2.840.10008.5.1.1.22	VOI LUT Box SOP Class	SOP Class	PS 3.4

1.2.840.10008.5.1.1.23	Presentation LUT SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.1.24	<i>Image Overlay Box SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.24.1	<i>Basic Print Image Overlay Box SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.25	<i>Print Queue SOP Instance (Retired)</i>	<i>Well-known Print Queue SOP Instance</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.26	<i>Print Queue Management SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.27	<i>Stored Print Storage SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.29	<i>Hardcopy Grayscale Image Storage SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.30	<i>Hardcopy Color Image Storage SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.31	<i>Pull Print Request SOP Class (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.32	<i>Pull Stored Print Management Meta SOP Class (Retired)</i>	<i>Meta SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.1.33	Media Creation Management SOP Class UID	SOP Class	PS3.4
1.2.840.10008.5.1.4.1.1.1	Computed Radiography Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.1	Digital X-Ray Image Storage – For Presentation	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.1.1	Digital X-Ray Image Storage – For Processing	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.2	Digital Mammography X-Ray Image Storage – For Presentation	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.2.1	Digital Mammography X-Ray Image Storage – For Processing	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.3	Digital Intra-oral X-Ray Image Storage – For Presentation	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.1.3.1	Digital Intra-oral X-Ray Image Storage – For Processing	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.2	CT Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.2.1	Enhanced CT Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.3	<i>Ultrasound Multi-frame Image Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.3.1	Ultrasound Multi-frame Image	SOP Class	PS 3.4

St

	Storage		
1.2.840.10008.5.1.4.1.1.4	MR Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.4.1	Enhanced MR Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.4.2	MR Spectroscopy Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.5</i>	<i>Nuclear Medicine Image Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.5.1.4.1.1.6</i>	<i>Ultrasound Image Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.6.1	Ultrasound Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.7	Secondary Capture Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.7.1	Multi-frame Single Bit Secondary Capture Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.7.2	Multi-frame Grayscale Byte Secondary Capture Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.7.3	Multi-frame Grayscale Word Secondary Capture Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.7.4	Multi-frame True Color Secondary Capture Image Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.8</i>	<i>Standalone Overlay Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.5.1.4.1.1.9</i>	<i>Standalone Curve Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.9.1.1	12-lead ECG Waveform Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.9.1.2	General ECG Waveform Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.9.1.3	Ambulatory ECG Waveform Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.9.2.1	Hemodynamic Waveform Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.9.3.1	Cardiac Electrophysiology Waveform Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.9.4.1	Basic Voice Audio Waveform Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.10</i>	<i>Standalone Modality LUT Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
<i>1.2.840.10008.5.1.4.1.1.11</i>	<i>Standalone VOI LUT Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.11.1	Grayscale Softcopy Presentation State Storage SOP Class	SOP Class	PS 3.4

1.2.840.10008.5.1.4.1.1.11.2	Color Softcopy Presentation State Storage SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.11.3	Pseudo-Color Softcopy Presentation State Storage SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.11.4	Blending Softcopy Presentation State Storage SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.12.1	X-Ray Angiographic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.12.1.1	Enhanced XA Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.12.2	X-Ray Radiofluoroscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.12.2.1	Enhanced XRF Image Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.12.3</i>	<i>X-Ray Angiographic Bi-Plane Image Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.20	Nuclear Medicine Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.66	Raw Data Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.66.1	Spatial Registration Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.66.2	Spatial Fiducials Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.66.3	Deformable Spatial Registration Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.66.4	Segmentation Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.67	Real World Value Mapping Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.77.1</i>	<i>VL Image Storage (Retired)</i>		
<i>1.2.840.10008.5.1.4.1.1.77.2</i>	<i>VL Multi-frame Image Storage (Retired)</i>		
1.2.840.10008.5.1.4.1.1.77.1.1	VL Endoscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.1.1	Video Endoscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.2	VL Microscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.2.1	Video Microscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.3	VL Slide-Coordinates Microscopic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.4	VL Photographic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.4.1	Video Photographic Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.5.1	Ophthalmic Photography 8 Bit Image Storage	SOP Class	PS 3.4

1.2.840.10008.5.1.4.1.1.77.1.5.2	Ophthalmic Photography 16 Bit Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.77.1.5.3	Stereometric Relationship Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.11	Basic Text SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.22	Enhanced SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.33	Comprehensive SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.40	Procedure Log Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.50	Mammography CAD SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.59	Key Object Selection Document	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.65	Chest CAD SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.88.67	X-Ray Radiation Dose SR	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.104.1	Encapsulated PDF Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.128	Positron Emission Tomography Image Storage	SOP Class	PS 3.4
<i>1.2.840.10008.5.1.4.1.1.129</i>	<i>Standalone PET Curve Storage (Retired)</i>	<i>SOP Class</i>	<i>PS 3.4</i>
1.2.840.10008.5.1.4.1.1.481.1	RT Image Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.2	RT Dose Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.3	RT Structure Set Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.4	RT Beams Treatment Record Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.5	RT Plan Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.6	RT Brachy Treatment Record Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.7	RT Treatment Summary Record Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.8	RT Ion Plan Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.1.481.9	RT Ion Beams Treatment Record Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.1.1	Patient Root Query/Retrieve Information Model – FIND	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.1.2	Patient Root Query/Retrieve Information Model – MOVE	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.1.3	Patient Root Query/Retrieve Information Model – GET	SOP Class	PS 3.4

1.2.840.10008.5.1.4.1.2.2.1	Study Root Query/Retrieve Information Model – FIND	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.2.2	Study Root Query/Retrieve Information Model – MOVE	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.2.3	Study Root Query/Retrieve Information Model – GET	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.3.1	<i>Patient/Study Only Query/Retrieve Information Model - FIND (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.3.2	<i>Patient/Study Only Query/Retrieve Information Model - MOVE (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.5.1.4.1.2.3.3	<i>Patient/Study Only Query/Retrieve Information Model - GET (Retired)</i>	SOP Class	PS 3.4
1.2.840.10008.5.1.4.31	Modality Worklist Information Model – FIND	SOP Class	PS 3.4
1.2.840.10008.5.1.4.32.1	General Purpose Worklist Information Model – FIND	SOP Class	PS 3.4
1.2.840.10008.5.1.4.32.2	General Purpose Scheduled Procedure Step SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.32.3	General Purpose Performed Procedure Step SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.32	General Purpose Worklist Management Meta SOP Class	Meta SOP Class	PS 3.4
1.2.840.10008.5.1.4.33	Instance Availability Notification SOP Class	SOP Class	PS 3.4
1.2.840.10008.5.1.4.37.1	General Relevant Patient Information Query	SOP Class	PS 3.4
1.2.840.10008.5.1.4.37.2	Breast Imaging Relevant Patient Information Query	SOP Class	PS 3.4
1.2.840.10008.5.1.4.37.3	Cardiac Relevant Patient Information Query	SOP Class	PS 3.4
1.2.840.10008.5.1.4.38.1	Hanging Protocol Storage	SOP Class	PS 3.4
1.2.840.10008.5.1.4.38.2	Hanging Protocol Information Model – FIND	SOP Class	PS 3.4
1.2.840.10008.5.1.4.38.3	Hanging Protocol Information Model – MOVE	SOP Class	PS 3.4
1.2.840.10008.15.0.3.1	dicomDeviceName	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.2	dicomDescription	LDAP OID	PS 3.15

1.2.840.10008.15.0.3.3	dicomManufacturer	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.4	dicomManufacturerModelName	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.5	dicomSoftwareVersion	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.6	dicomVendorData	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.7	dicomAETitle	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.8	dicomNetworkConnectionReference	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.9	dicomApplicationCluster	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.10	dicomAssociationInitiator	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.11	dicomAssociationAcceptor	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.12	dicomHostname	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.13	dicomPort	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.14	dicomSOPClass	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.15	dicomTransferRole	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.16	dicomTransferSyntax	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.17	dicomPrimaryDeviceType	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.18	dicomRelatedDeviceReference	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.19	dicomPreferredCalledAETitle	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.20	dicomTLSCyphersuite	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.21	dicomAuthorizedNodeCertificateReference	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.22	dicomThisNodeCertificateReference	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.23	dicomInstalled	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.24	dicomStationName	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.25	dicomDeviceSerialNumber	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.26	dicomInstitutionName	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.27	dicomInstitutionAddress	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.28	dicomInstitutionDepartmentName	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.29	dicomIssuerOfPatientID	LDAP OID	PS 3.15
1.2.840.10008.15.0.3.30	dicomPreferredCallingAETitle	LDAP OID	PS 3.15



1.2.840.10008.15.0.3.31	dicomSupportedCharacterSet	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.1	dicomConfigurationRoot	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.2	dicomDevicesRoot	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.3	dicomUniqueAETitlesRegistryRoot	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.4	dicomDevice	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.5	dicomNetworkAE	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.6	dicomNetworkConnection	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.7	dicomUniqueAETitle	LDAP OID	PS 3.15
1.2.840.10008.15.0.4.8	dicomTransferCapability	LDAP OID	PS 3.15

**Table A-2**  
**Well-known Frames of Reference**

UID Value	UID NAME	Normative Reference
1.2.840.10008.1.4.1.1	Talairach Brain Atlas Frame of Reference	Talairach J. and Tournoux P. <i>Co-Planar stereotactic atlas of the human brain</i> . Stuttgart: Georg Thieme Verlag, 1988.
1.2.840.10008.1.4.1.2	SPM2 T1 Frame of Reference	SPM2 /templates/T1.mnc
1.2.840.10008.1.4.1.3	SPM2 T2 Frame of Reference	SPM2 /templates/T2.mnc
1.2.840.10008.1.4.1.4	SPM2 PD Frame of Reference	SPM2 /templates/PD.mnc
1.2.840.10008.1.4.1.5	SPM2 EPI Frame of Reference	SPM2 /templates/EPI.mnc
1.2.840.10008.1.4.1.6	SPM2 FIL T1 Frame of Reference	SPM2 /templates/filt1.mnc
1.2.840.10008.1.4.1.7	SPM2 PET Frame of Reference	SPM2 /templates/PET.mnc
1.2.840.10008.1.4.1.8	SPM2 TRANSM Frame of Reference	SPM2 /templates/Transm.mnc
1.2.840.10008.1.4.1.9	SPM2 SPECT Frame of Reference	SPM2 /templates/SPECT.mnc
1.2.840.10008.1.4.1.10	SPM2 GRAY Frame of Reference	SPM2 /apriori/gray.mnc
1.2.840.10008.1.4.1.11	SPM2 WHITE Frame of Reference	SPM2 /apriori/white.mnc
1.2.840.10008.1.4.1.12	SPM2 CSF Frame of Reference	SPM2 /apriori/csf.mnc
1.2.840.10008.1.4.1.13	SPM2 BRAINMASK	SPM2 /apriori/brainmask.mnc

	Frame of Reference	
1.2.840.10008.1.4.1.14	SPM2 AVG305T1 Frame of Reference	SPM2 /canonical/avg305T1.mnc
1.2.840.10008.1.4.1.15	SPM2 AVG152T1 Frame of Reference	SPM2 /canonical/avg152T1.mnc
1.2.840.10008.1.4.1.16	SPM2 AVG152T2 Frame of Reference	SPM2 /canonical/avg152T2.mnc
1.2.840.10008.1.4.1.17	SPM2 AVG152PD Frame of Reference	SPM2 /canonical/avg152PD.mnc
1.2.840.10008.1.4.1.18	SPM2 SINGLESUBJT1 Frame of Reference	SPM2 /canonical/single_subj_T1.mnc
1.2.840.10008.1.4.2.1	ICBM 452 T1 Frame of Reference	ICBM452 T1 Atlas
1.2.840.10008.1.4.2.2	ICBM Single Subject MRI Frame of Reference	ICBM Single Subject MRI Anatomical Template

SPM2 (Statistical Parametric Mapping) templates are available at <http://www.fil.ion.ucl.ac.uk/~spm/>, and they are described at <http://www.fil.ion.ucl.ac.uk/~spm/templates/>.

ICBM templates are available at [http://www.loni.ucla.edu/ICBM/ICBM\\_ICBMAtlases.html](http://www.loni.ucla.edu/ICBM/ICBM_ICBMAtlases.html).